

RBP4 (Human) - EIA Kit - Extraction Free*

Catalog#	Standard size	Price	Order
EK-028-28	1 kit	\$495.00	<input type="button" value="ORDER"/>

Product description

More information

References

Retinol Binding Protein 4

RBP-4

Protocol [Please Inquire \(email\)](#) - **No Sample Extraction Required***

*Serum and plasma do not require extraction. All other sample types may need extraction.

Note: Read this protocol thoroughly prior to performing the assay!

Preliminary tests must be done to determine the dilution factor of samples for optimal results. Un-extracted human plasma sample of normal subject should be diluted to 1:1000 - 1:5000 using 1x assay buffer. It is up to the discretion of each laboratory to determine appropriate conditions for this assay.

Sample Preparation
[Blood collection](#)
[Tissue extraction](#)
[CSF extraction](#)

Catalog # EK-028-28

Standard Size 96 wells

Sensitivity Minimum Detectable Concentration = 3.12ng/ml

Range 0.1-1000ng/ml

Linear Range 3.12-31.4ng/ml

Cross Reactivity	Peptide	%Cross-reactivity
	RBP-4(H)	100
	RBP-4(R)	0
	Adiponectin (H)	0
	Resistin (H)	0
	Apelin 36 (H)	0
	Visfatin (400-450)(H)	0
	Insulin(H)	0
	Leptin (Human)	0
	Visfatin C-terminal (Human)	0
	TNF-alpha (Human)	0

Standard
Curve Image

