



## Bull's Eye Decloaker, 20X

Pretreatment Reagent

Control Number: 902-1000B-120412

**Catalog Number:** BRR 1000B MX

**Description:** 500 ml; concentrate

**Intended Use:**

For Research Use Only. Not for use in diagnostic procedures.

**Summary & Explanation:**

Bull's Eye is a heat retrieval solution that allows the end-user to block for endogenous peroxidase and block for non-specific background staining while simultaneously antigen retrieving tissues. Bull's Eye is a universal buffer compatible with virtually all antibodies and thus eliminates the need for EDTA or high pH retrieval buffers. Bull's Eye can be used with Biocare's digital electric pressure cooker (Decloaking Chamber), a steamer, water bath or microwave oven. Antibody titers are doubled and tripled when compared to citrate buffer, pH 6.0. Bull's-Eye incorporates Assure Technology™, a color-coded high temperature pH indicator solution. The end-user is assured by visual inspection that the solution is at the correct dilution and pH. This product is specially formulated for superior pH stability at high temperatures and will help prevent the possibility of losing pH sensitive antigens. Bull's Eye is non-toxic, non-flammable, odorless and sodium azide and thimerosal free.

**Known Applications:**

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

**Supplied As:**

**500ml**

Bull's Eye Decloaker, 20X concentrate (BRR1000BMX)

**Materials and Reagents Needed But Not Provided:**

- Microscope slides, positively charged
- Desert Chamber\* (Drying oven)
- Positive and negative tissue controls
- Xylene (Could be substituted with xylene substitute\*)
- Ethanol or reagent alcohol
- Decloaking Chamber\* (Pressure cooker)
- Deionized or distilled water
- Wash buffer\*(TBS/PBS)
- Enzyme digestion\*
- Avidin-Biotin Blocking Kit\*(Labeled Streptavidin Kits Only)
- Peroxidase block\*
- Protein block\*
- Primary antibody\*
- Negative control reagents\*
- Detection kits\*
- Detection components\*
- Chromogens\*
- Hematoxylin\*
- Bluing reagent\*
- Mounting medium\*

\* Biocare Medical Products: Refer to a Biocare Medical catalog for further information regarding catalog numbers and ordering information. Certain reagents listed above are based on specific application and detection system used.

**Storage and Stability:**

Store at room temperature. Do not use after expiration date printed on vial. If reagents are stored under conditions other than those specified in the package insert, they must be verified by the user. Diluted reagents should be used promptly; any remaining reagent should be stored at room temperature.

**Protocol Recommendations:**

1. Deparaffinize tissues and hydrate to water. If necessary, block for endogenous peroxidase and wash in DI water.
2. Dilute concentrated Bull's Eye Decloaker at a ratio of 1:20 (1 ml Bull's Eye to 19 ml of deionized water).
3. Place slides into 1X retrieval solution in a slide container (e.g. Coplin jar, Tissue -Tek™ staining dish or metal slide canister).
4. Retrieve sections under pressure using Biocare's Decloaking Chamber. Follow the recommendations on the antibody data sheet and Table 1 (below).
5. Check solution for appropriate color change. (See Technical Note #1)
6. Gently rinse by gradually adding DI water to the solution, then remove slides and rinse with DI water.

**Technical Notes:**

1. Concentrated Bull's Eye Decloaker is a bright yellow color. RTU or 1X solution is a pale yellow color. When the solution reaches 80-125°C, the solution turns yellow and indicates that the high temperature solution is at correct pH. Should the pH rise above 7.0, the solution turns a fuschia red color. Should the pH drop too low, the solution turns a pink color.
2. Bull's Eye Decloaker may be used with various heat retrieval methods including a microwave oven, pressure cooker, hot water bath or steamer.
3. If using Biocare's Desert Chamber Pro (a programmable turbo-action drying oven), dry sections at 37°C for 30-60 minutes and then dry slides at 60°C for 30 minutes.
4. Use positive charged slides (use Biocare's Kling-On HIER Slides) and cut tissues at 4-5 microns. Do not use any adhesives in the water bath. Poor fixation and processing of tissues will cause tissue sections to fall off the slides, especially fatty tissues such as breast. Tissues should be fixed a minimum of 6-12 hours.

**Table 1. Biocare Decloaking Chamber Protocol Equivalence**

Protocol	DC2002/DC2008	DC NxGen
125°C/30 sec	125°C/30 sec	110°C/10 to 15 min
95°C/40 min	95°C/40 min	95°C/40 min
90°C/15 min	90°C/15 min	90°C/15 min
80°C/60 min	80°C/60 min	80°C/60 min
60°C/15 hrs	60°C/15 hrs	60°C/15 hrs

**Limitations:**

This product is provided for Research Use Only (RUO) and is not for use in diagnostic procedures. Suitability for specific applications may vary and it is the responsibility of the end user to determine the appropriate application for its use.

**Precautions:**

1. This product is not classified as hazardous. The preservative used in this reagent is Proclin 300 and the concentration is less than 0.25%. Overexposure to Proclin 300 can cause skin and eye irritation and irritation to mucous membranes and upper respiratory tract. The concentration of Proclin 300 in this product does not meet the OSHA criteria for a hazardous substance. Wear disposable gloves when handling reagents.
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water.
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the vial.

**Bull's Eye Decloaker, 20X**

Pretreatment Reagent

Control Number: 902-1000B-120412

**Precautions Cont'd:**

6. The MSDS is available upon request and is located at <http://biocare.net/support/msds/>.

7. Consult OSHA, federal, state or local regulations for disposal of any toxic substances. Proclin™ is a trademark of Rohm and Haas Company, or of its subsidiaries or affiliates.

**Technical Support:**

Contact Biocare's Technical Support at 1-800-542-2002 for questions regarding this product.

**Warranty:**

There are no warranties, expressed or implied, which extend beyond this description. Biocare is not liable for property damage, personal injury, or economic loss caused by this product.