

Nuclear Decloaker, 10X

Pretreatment Reagent
Control Number: 902-911-022912

ISO 9001&13485 CERTIFIED

Catalog Number: BRR 911 M

Description: 500 ml; concentrate

Intended Use:

For Research Use Only. Not for use in diagnostic procedures

Summary & Explanation:

Nuclear Decloaker is a Tris buffer heat retrieval solution, pH 9.5. It is specially designed for nuclear antigens including ER, PR, Ki-67, p53, cyclin D1, TdT and TTF -1. Other antibodies that show improved staining include ALK1, ALKc, bcl-2, CA125, CD3 (M), CD15, CD30, CD31, CD79a, MART-1, Tyrosinase and Villin. Nuclear Decloaker can be used with Biocare's digital electric pressure cooker (Decloaking Chamber), a steamer, a water bath, or a microwave oven. Nuclear Decloaker incorporates Assure™ technology, a color-coded high temperature pH indicator solution. The end-user is assured by visual inspection that the solution is at the correct dilution and pH. This product is specially formulated for superior pH stability at high temperatures and will help prevent the possibility of losing pH sensitive antigens. Nuclear Decloaker is non-toxic, non-flammable, odorless and sodium azide and thimerosal free

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As:

500ml

Nuclear Decloaker, 10X concentrate (BRR911M)

Materials and Reagents Needed But Not Provided):

Microscope slides, positively charged

Desert Chamber* (Drying oven)

Positive and negative tissue controls

Xylene (Could be substituted with xylene substitute*)

Ethanol or reagent alcohol

Decloaking Chamber* (Pressure cooker)

Deionized or distilled water

Wash buffer*(TBS/PBS)

Enzyme digestion*

Avidin-Biotin Blocking Kit*(Labeled Streptavidin Kits Only)

Peroxidase block*

Protein block*

Primary antibody*

Negative control reagents*

Detection kits*

Detection components*

Chromogens*

Hematoxylin*

Bluing reagent*

Mounting medium*

* Biocare Medical Products: Refer to a Biocare Medical catalog for further information regarding catalog numbers and ordering information. Certain reagents listed above are based on specific application and detection system used.

Storage and Stability:

Store at room temperature. Do not use after expiration date printed on vial. If reagents are stored under conditions other than those specified in the package insert, they must be verified by the user. Diluted reagents should be used promptly; any remaining reagent should be stored at room temperature.

Protocol Recommendations:

- Deparaffinize tissues and hydrate to water. If necessary, block for endogenous peroxidase and wash in DI water.
- Dilute concentrated Nuclear Decloaker at a ratio of 1:10 (1 ml Nuclear Decloaker to 9 ml deionized water)
- 3. Place slides into 1X retrieval solution in a slide container (e.g. Coplin jar, Tissue-TekTM staining dish or metal slide canister).
- 4. Retrieve sections under pressure using Biocare's Decloaking Chamber. Follow the recommendations on the antibody data sheet and Table 1 (below).
- 5. Check solution for appropriate color change. (See Technical Note #1)
- 6. Gently rinse by gradually adding DI water to the solution, then remove slides and rinse with DI water.

Technical Notes:

- Concentrated Nuclear Decloaker is a bright fuschia red color. RTU or 1X solution is
 pale fuschia red color. When the solution reaches 80-125°C, the solution stays
 fuschia red and indicates that the high temperature solution is at correct pH. If the
 solution turns orange or yellow, the pH of the buffer is incorrect.
- Nuclear Decloaker may be used with various heat retrieval methods including a microwave oven, pressure cooker, hot water bath or steamer.
- 3. If using Biocare's Desert Chamber Pro (a programmable turbo-action drying oven), dry sections at 25°C overnight or at 37°C for 30-60 minutes and then dry slides at 60°C for 30 minutes.
- 4. Use positive charged slides (use Biocare's Kling-On HIER Slides) and cut tissues at 4-5 microns. So not use any adhesives in the water bath. Poor fixation and processing of tissues will cause tissue sections to fall off the slides, especially fatty tissues such as breast. Tissues should be fixed a minimum of 6-12 hours.

Table 1. Biocare Decloaking Chamber Protocol Equivalence

<u>Protocol</u>	DC2002/DC2008	DC NxGen
125°C/30 sec	125°C/30 sec	110°C/10 to 15 min
95°C/40 min	95°C/40 min	95°C/40 min
90°C/15 min	90°C/15 min	90°C/15 min
80°C/60 min	80°C/60 min	80°C/60 min
60°C/15 hrs	60°C/15 hrs	60°C/15 hrs

Performance Characteristics:

The protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

Ouality Control:

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2). CLSI Wayne, PA, USA (www.clsi.org). 2011

tel: 800-799-9499 fax: 925-603-8080 www.biocare.net



continually researching the most cost-effective and quality-aware diagnostic solutions



Nuclear Decloaker, 10X

Pretreatment Reagent
Control Number: 902-911-022912

ISO 9001&13485 CERTIFIED

Precautions:

This product is not classified as hazardous. The preservative used in this reagent is Proclin 950 and the concentration is less than 0.25%. Overexposure to Proclin 950 can cause skin and eye irritation and irritation to mucous membranes and upper respiratory tract. The concentration of Proclin 950 in this product does not meet the OSHA criteria for a hazardous substance. Wear disposable gloves when handling reagents. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. Microbial contamination of reagents may result in an increase in nonspecific staining. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change. The MSDS is available upon request and is located at http://biocare.net/support/msds/.

Consult OSHA, federal, state or local regulations for disposal of any toxic substances. ProclinTM is a trademark of Rohm and Haas Company, or of its subsidiaries or affiliates.

Troubleshooting:

Follow the reagent specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542 -2002.

Limitations & Warranty:

There are no warranties, expressed or implied, which extend beyond this description. Biocare is not liable for property damage, personal injury, or economic loss caused by this product.