



Estriol ELISA Kit

Item No. 582281

www.caymanchem.com

Customer Service 800.364.9897

Technical Support 888.526.5351

1180 E. Ellsworth Rd · Ann Arbor, MI · USA

TABLE OF CONTENTS

GENERAL INFORMATION	3	Materials Supplied
	4	Safety Data
	4	Precautions
	5	If You Have Problems
	5	Storage and Stability
	5	Materials Needed but Not Supplied
INTRODUCTION	6	Background
	6	About This Assay
	7	Description of AChE Competitive ELISAs
	9	Biochemistry of Acetylcholinesterase
	11	Definition of Key Terms
PRE-ASSAY PREPARATION	12	Buffer Preparation
	13	Sample Preparation
ASSAY PROTOCOL	16	Preparation of Assay-Specific Reagents
	19	Plate Set Up
	20	Performing the Assay
ANALYSIS	23	Calculations
	25	Performance Characteristics
RESOURCES	32	Troubleshooting
	33	References
	34	Plate Template
	35	Notes

GENERAL INFORMATION

Materials Supplied

Item Number	Item	96 wells Quantity/Size	480 wells Quantity/Size
482282	Estriol ELISA Antiserum	1 vial/100 dtn	1 vial/500 dtn
482280	Estriol AChE Tracer	1 vial/100 dtn	1 vial/500 dtn
482284	Estriol ELISA Standard	1 vial	1 vial
400060	ELISA Buffer Concentrate (10X)	2 vials/10 ml	4 vials/10 ml
400062	Wash Buffer Concentrate (400X)	1 vial/5 ml	1 vial/12.5 ml
400035	Polysorbate 20	1 vial/3 ml	1 vial/3 ml
400004/400006	Mouse Anti-Rabbit IgG Coated Plate	1 plate	5 plates
400012	96-Well Cover Sheet	1 cover	5 covers
400050	Ellman's Reagent	3 vials/100 dtn	6 vials/250 dtn
400040	ELISA Tracer Dye	1 vial	1 vial
400042	ELISA Antiserum Dye	1 vial	1 vial

If any of the items listed above are damaged or missing, please contact our Customer Service department at (800) 364-9897 or (734) 971-3335. We cannot accept any returns without prior authorization.



WARNING: THIS PRODUCT IS FOR RESEARCH ONLY - NOT FOR HUMAN OR VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

Safety Data

This material should be considered hazardous until further information becomes available. Do not ingest, inhale, get in eyes, on skin, or on clothing. Wash thoroughly after handling. Before use, the user must review the complete Safety Data Sheet, which has been sent *via* email to your institution.

Precautions

Please read these instructions carefully before beginning this assay.

The reagents in this kit have been tested and formulated to work exclusively with Cayman Chemical's AChE ELISA Kits. This kit may not perform as described if any reagent or procedure is replaced or modified.

When compared to quantification by LC/MS or GC/MS, it is not uncommon for immunoassays to report higher analyte concentrations. While LC/MS or GC/MS analyses typically measure only a single compound, antibodies used in immunoassays sometimes recognize not only the target molecule, but also structurally related molecules, including biologically relevant metabolites. In many cases, measurement of both the parent molecule and metabolites is more representative of the overall biological response than is the measurement of a short-lived parent molecule. It is the responsibility of the researcher to understand the limits of both assay systems and to interpret their data accordingly.

If You Have Problems

Technical Service Contact Information

Phone: 888-526-5351 (USA and Canada only) or 734-975-3888
Fax: 734-971-3641
Email: techserv@caymanchem.com
Hours: M-F 8:00 AM to 5:30 PM EST

In order for our staff to assist you quickly and efficiently, please be ready to supply the lot number of the kit (found on the outside of the box).

Storage and Stability

This kit will perform as specified if stored as directed at -20°C and used before the expiration date indicated on the outside of the box.

Materials Needed But Not Supplied

1. A plate reader capable of measuring absorbance between 405-420 nm.
2. Adjustable pipettes and a repeating pipettor.
3. A source of 'UltraPure' water. Water used to prepare all ELISA reagents and buffers must be deionized and free of trace organic contaminants ('UltraPure'). Use activated carbon filter cartridges or other organic scavengers. Glass distilled water (even if double distilled), HPLC-grade water, and sterile water (for injections) are not adequate for ELISA. *NOTE: UltraPure water is available for purchase from Cayman (Item No. 400000).*
4. Materials used for Sample Preparation (see page 13).

Background

Estriol, a metabolite of estradiol and a major estrogen produced during pregnancy, has historically been used as an indicator of fetal well-being. In the final weeks before parturition, estriol levels increase significantly. In a longitudinal study in healthy pregnant women, total plasma estriol levels increased from <10 ng/ml at 8-10 weeks of gestation to approximately 150 ng/ml at week 38.¹ The majority of the estriol synthesized in the later stages of pregnancy originates from fetal dehydroepiandrosterone sulfate (DHEAS) and serves as a direct marker of fetal adrenal gland activity.² In the 1960's and 1970's plasma and urinary estriol levels were used as biological fluids for measuring estriol. However, saliva contains primarily unbound and unconjugated estriol, the biologically active form of the hormone, and therefore is a commonly used biological fluid for monitoring estriol levels.^{2,3} Plasma levels of estriol in males and non-pregnant females is less than 15 pg/ml.

About This Assay

Cayman's Estriol ELISA Kit is a competitive assay that can be used for quantification of estriol in plasma and other sample matrices. In an overnight incubation at 4°C, the assay has a range from 0.9-2,000 pg/ml and a sensitivity (80% B/B₀) of approximately 4 pg/ml. In a two hour room temperature format, the assay also has a range from 0.9-2,000 pg/ml but a 80% B/B₀ value of approximately 12 pg/ml.

Description of AChE Competitive ELISAs^{4,5}

This assay is based on the competition between estriol and an estriol-acetylcholinesterase (AChE) conjugate (estriol tracer) for a limited number of estriol-specific rabbit antiserum binding sites. Because the concentration of the estriol tracer is held constant while the concentration of estriol varies, the amount of estriol tracer that is able to bind to the rabbit antiserum will be inversely proportional to the concentration of estriol in the well. This rabbit antiserum-estriol (either free or tracer) complex binds to the mouse monoclonal anti-rabbit IgG that has been previously attached to the well. The plate is washed to remove any unbound reagents and then Ellman's Reagent (which contains the substrate to AChE) is added to the well. The product of this enzymatic reaction has a distinct yellow color and absorbs strongly at 412 nm. The intensity of this color, determined spectrophotometrically, is proportional to the amount of estriol tracer bound to the well, which is inversely proportional to the amount of free estriol present in the well during the incubation; or

$$\text{Absorbance} \propto [\text{Bound Estriol Tracer}] \propto 1/[\text{Estriol}]$$

A schematic of this process is shown in Figure 1, on page 8.




Figure 1. Schematic of the AChE ELISA

Biochemistry of Acetylcholinesterase

The electric organ of the electric eel, *E. electricus*, contains an avid AChE capable of massive catalytic turnover during the generation of its electrochemical discharges. The electric eel AChE has a clover leaf-shaped tertiary structure consisting of a triad of tetramers attached to a collagen-like structural fibril. This stable enzyme is capable of high turnover ($64,000 \text{ s}^{-1}$) for the hydrolysis of acetylthiocholine.

A molecule of the analyte covalently attached to a molecule of AChE serves as the tracer in ACE™ enzyme immunoassays. Quantification of the tracer is achieved by measuring its AChE activity with Ellman's Reagent. This reagent consists of acetylthiocholine and 5,5'-dithio-bis-(2-nitrobenzoic acid). Hydrolysis of acetylthiocholine by AChE produces thiocholine (see Figure 2, on page 10). The non-enzymatic reaction of thiocholine with 5,5'-dithio-bis-(2-nitrobenzoic acid) produces 5-thio-2-nitrobenzoic acid, which has a strong absorbance at 412 nm ($\epsilon = 13,600$).

AChE has several advantages over other enzymes commonly used for enzyme immunoassays. Unlike horseradish peroxidase, AChE does not self-inactivate during turnover. This property of AChE also allows redevelopment of the assay if it is accidentally splashed or spilled. In addition, the enzyme is highly stable under the assay conditions, has a wide pH range (pH 5-10), and is not inhibited by common buffer salts or preservatives. Since AChE is stable during the development step, it is unnecessary to use a 'stop' reagent, and the plate may be read whenever it is convenient.




Figure 2. Reaction catalyzed by acetylcholinesterase

Definition of Key Terms

Blank: background absorbance caused by Ellman's Reagent. The blank absorbance should be subtracted from the absorbance readings of all the other wells, including NSB wells.

Total Activity: total enzymatic activity of the AChE-linked tracer. This is analogous to the specific activity of a radioactive tracer.

NSB (Non-Specific Binding): non-immunological binding of the tracer to the well. Even in the absence of specific antibody a very small amount of tracer still binds to the well; the NSB is a measure of this low binding. Do not forget to subtract the Blank absorbance values.

B_0 (Maximum Binding): maximum amount of the tracer that the antibody can bind in the absence of free analyte.

% B/B_0 (%Bound/Maximum Bound): ratio of the absorbance of a particular sample or standard well to that of the maximum binding (B_0) well.

Standard Curve: a plot of the % B/B_0 values *versus* concentration of a series of wells containing various known amounts of analyte.

Dtn: determination, where one dtn is the amount of reagent used per well.

Cross Reactivity: numerical representation of the relative reactivity of this assay towards structurally related molecules as compared to the primary analyte of interest. Biomolecules that possess similar epitopes to the analyte can compete with the assay tracer for binding to the primary antibody. Substances that are superior to the analyte in displacing the tracer result in a cross reactivity that is greater than 100%. Substances that are inferior to the primary analyte in displacing the tracer result in a cross reactivity that is less than 100%. Cross reactivity is calculated by comparing the mid-point (50% B/B_0) value of the tested molecule to the mid-point (50% B/B_0) value of the primary analyte when each is measured in assay buffer using the following formula:

$$\% \text{ Cross Reactivity} = \left[\frac{50\% B/B_0 \text{ value for the primary analyte}}{50\% B/B_0 \text{ value for the potential cross reactant}} \right] \times 100\%$$

PRE-ASSAY PREPARATION

NOTE: Water used to prepare all ELISA reagents and buffers must be deionized and free of trace organic contaminants ('UltraPure'). Use activated carbon filter cartridges or other organic scavengers. Glass distilled water (even if double distilled), HPLC-grade water, and sterile water (for injections) are not adequate for ELISA. UltraPure water may be purchased from Cayman (Item No. 400000).

Buffer Preparation

Store all diluted buffers at 4°C; they will be stable for about two months.

1. ELISA Buffer Preparation

Dilute the contents of one vial of ELISA Buffer Concentrate (10X) (Item No. 400060) with 90 ml of UltraPure water. Be certain to rinse the vial to remove any salts that may have precipitated. *NOTE: It is normal for the concentrated buffer to contain crystalline salts after thawing. These will completely dissolve upon dilution with water.*

2. Wash Buffer Preparation

5 ml vial Wash Buffer Concentrate (400X) (96-well kit; Item No. 400062): Dilute to a total volume of 2 liters with UltraPure water and add 1 ml of Polysorbate 20 (Item No. 400035).

OR

12.5 ml vial Wash Buffer Concentrate (400X) (480-well kit; Item No. 400062): Dilute to a total volume of 5 liters with UltraPure water and add 2.5 ml of Polysorbate 20 (Item No. 400035).

Smaller volumes of Wash Buffer can be prepared by diluting the Wash Buffer Concentrate 1:400 and adding Polysorbate 20 (0.5 ml/liter of Wash Buffer).

NOTE: Polysorbate 20 is a viscous liquid and cannot be measured by a regular pipette. A positive displacement pipette or a syringe should be used to deliver small quantities accurately.

Sample Preparation

In general, urine and tissue culture supernatant samples may be diluted with ELISA Buffer and added directly to the assay well. Plasma, serum, whole blood, as well as other heterogeneous mixtures such as lavage fluids and aspirates often contain contaminants which can interfere in the assay. It is best to check for interference before embarking on a large number of sample measurements. To test for interference, dilute one or two test samples to obtain at least two different dilutions of each sample between approximately 20 and 1,000 pg/ml (*i.e.*, to be within 20-80% B/B₀). If the two different dilutions of the sample show good correlation (differ by 20% or less) in the final calculated estriol concentration, purification is not required. If you do not see good correlation of the different dilutions, purification is advised. The Extraction Protocol below is one such method.

General Precautions

- All samples must be free of organic solvents prior to assay.
- Samples should be assayed immediately after collection; samples that cannot be assayed immediately should be stored at -80°C.
- Samples of rabbit origin may contain antibodies which interfere with the assay by binding to the mouse anti-rabbit plate. We recommend that all rabbit samples be purified prior to use in this assay.

Extraction Protocol

The protocol described below is a suggestion only. You may choose a different protocol based on your own requirements, sample type, and expertise.

NOTE: We do not recommend the use of plastic vials or caps for this procedure. The ethyl acetate/hexane procedure may extract interfering compounds from the plastic.

1. Aliquot an equal amount of each sample into clean test tubes (500 μ l is recommended). If your samples need to be concentrated, a larger volume should be used.
2. Add four times the sample volume with ethyl acetate/hexane (50:50) and vortex 2 x 10 seconds. Allow the layers to separate. Transfer the ethyl acetate/hexane layer (upper layer) to a clean test tube. Repeat this step three times.
3. Evaporate the ethyl acetate/hexane by vacuum centrifugation or by heating to 30°C under a gentle stream of nitrogen.
4. Reconstitute the extract in a volume of ELISA Buffer equivalent to the original sample volume. Dilute samples can be concentrated by reconstituting the extract in a volume smaller than that of the original sample volume (e.g., an original sample volume of 5 ml reconstituted in 500 μ l will be concentrated by a factor of 10).




Figure 3. Plasma Validation

Plasma samples were spiked with estriol, extracted as described in the **Sample Preparation** section (see page 13) and analyzed using the Estriol ELISA Kit. The y-intercept corresponds to the amount of estriol in unspiked plasma. Error bars represent standard deviations obtained from multiple dilutions of each sample.

ASSAY PROTOCOL

Preparation of Assay-Specific Reagents

Estriol ELISA Standard

Equilibrate a pipette tip in ethanol by repeatedly filling and expelling the tip with ethanol several times. Using the equilibrated pipette tip, transfer 100 μ l of the Estriol ELISA Standard (Item No. 482284) into a clean test tube, then dilute with 900 μ l UltraPure water. The concentration of this solution (the bulk standard) will be 20 ng/ml.

NOTE: If assaying culture media samples that have not been diluted with ELISA Buffer, culture medium should be used in place of ELISA Buffer for dilution of the standard curve.

To prepare the standard for use in ELISA: Obtain eight clean test tubes and number them #1 through #8. Aliquot 900 μ l ELISA Buffer to tube #1 and 600 μ l ELISA Buffer to tubes #2-8. Transfer 100 μ l of the bulk standard (20 ng/ml) to tube #1 and mix thoroughly. Serially dilute the standard by removing 300 μ l from tube #1 and placing in tube #2; mix thoroughly. Next, remove 300 μ l from tube #2 and place it into tube #3; mix thoroughly. Repeat this process for tubes #4-8. These diluted standards should not be stored for more than 24 hours.




Figure 4. Preparation of the estriol standards

Estriol AChE Tracer

Reconstitute the Estriol AChE Tracer as follows:

100 dtn Estriol AChE Tracer (96-well kit; Item No. 482280): Reconstitute with 6 ml ELISA Buffer.

OR

500 dtn Estriol AChE Tracer (480-well kit; Item No. 482280): Reconstitute with 30 ml ELISA Buffer.

Store the reconstituted Estriol AChE Tracer at 4°C (*do not freeze!*) and use within four weeks. A 20% surplus of tracer has been included to account for any incidental losses.

Tracer Dye Instructions (optional)

This dye may be added to the tracer, if desired, to aid in visualization of tracer-containing wells. Add the dye to the reconstituted tracer at a final dilution of 1:100 (add 60 μ l of dye to 6 ml tracer or add 300 μ l of dye to 30 ml of tracer). *NOTE: Do not store tracer with dye for more than 24 hours.*

Estriol ELISA Antiserum

Reconstitute the Estriol ELISA Antiserum as follows:

100 dtn Estriol ELISA Antiserum (96-well kit; Item No. 482282):
Reconstitute with 6 ml ELISA Buffer.

OR

500 dtn Estriol ELISA Antiserum (480-well kit; Item No. 482282):
Reconstitute with 30 ml ELISA Buffer.

Store the reconstituted Estriol ELISA Antiserum at 4°C. It will be stable for at least four weeks. A 20% surplus of antiserum has been included to account for any incidental losses.

Antiserum Dye Instructions (optional)

This dye may be added to the antiserum, if desired, to aid in visualization of antiserum-containing wells. Add the dye to the reconstituted antiserum at a final dilution of 1:100 (add 60 µl of dye to 6 ml antiserum or add 300 µl of dye to 30 ml of antiserum). *NOTE: Do not store antiserum with dye for more than 24 hours.*

Plate Set Up

The 96-well plate(s) included with this kit is supplied ready to use. It is not necessary to rinse the plate(s) prior to adding the reagents. *NOTE: If you do not need to use all the strips at once, place the unused strips back in the plate packet and store at 4°C. Be sure the packet is sealed with the desiccant inside.*

Each plate or set of strips must contain a minimum of two blanks (Blk), two non-specific binding wells (NSB), two maximum binding wells (B_0), and an eight point standard curve run in duplicate. *NOTE: Each assay must contain this minimum configuration in order to ensure accurate and reproducible results.* Each sample should be assayed at two dilutions and each dilution should be assayed in duplicate. For statistical purposes, we recommend assaying samples in triplicate.

A suggested plate format is shown below in Figure 5. The user may vary the location and type of wells present as necessary for each particular experiment. The plate format provided below has been designed to allow for easy data analysis using a convenient spreadsheet offered by Cayman Chemical (see **Analysis**, page 23, for more details). We suggest you record the contents of each well on the template sheet provided (see page 34).

	1	2	3	4	5	6	7	8	9	10	11	12
A	Blk	S1	S1	1	1	1	9	9	9	17	17	17
B	Blk	S2	S2	2	2	2	10	10	10	18	18	18
C	NSB	S3	S3	3	3	3	11	11	11	19	19	19
D	NSB	S4	S4	4	4	4	12	12	12	20	20	20
E	B_0	S5	S5	5	5	5	13	13	13	21	21	21
F	B_0	S6	S6	6	6	6	14	14	14	22	22	22
G	B_0	S7	S7	7	7	7	15	15	15	23	23	23
H	TA	S8	S8	8	8	8	16	16	16	24	24	24

Blk - Blank
TA - Total Activity
NSB - Non-Specific Binding
 B_0 - Maximum Binding
S1-S8 - Standards 1-8
1-24 - Samples

Figure 5. Sample plate format

Performing the Assay

Pipetting Hints

- Use different tips to pipette each reagent.
- Before pipetting each reagent, equilibrate the pipette tip in that reagent (*i.e.*, slowly fill the tip and gently expel the contents, repeat several times).
- Do not expose the pipette tip to the reagent(s) already in the well.

Addition of the Reagents

1. ELISA Buffer

Add 100 μ l ELISA Buffer to NSB wells. Add 50 μ l ELISA Buffer to B₀ wells. If culture medium was used to dilute the standard curve, substitute 50 μ l of culture medium for ELISA Buffer in the NSB and B₀ wells (*i.e.*, add 50 μ l culture medium to NSB and B₀ wells and 50 μ l ELISA Buffer to NSB wells).

2. Estriol ELISA Standard

Add 50 μ l from tube #8 to both of the lowest standard wells (S8). Add 50 μ l from tube #7 to each of the next two standard wells (S7). Continue with this procedure until all the standards are aliquoted. The same pipette tip should be used to aliquot all the standards. Before pipetting each standard, be sure to equilibrate the pipette tip in that standard.

3. Samples

Add 50 μ l of sample per well. Each sample should be assayed at a minimum of two dilutions. Each dilution should be assayed in duplicate (triplicate recommended).

4. Estriol AChE Tracer

Add 50 μ l to each well *except* the TA and the Blk wells.

5. Estriol ELISA Antiserum

Add 50 μ l to each well *except* the TA, the NSB, and the Blk wells.

Well	ELISA Buffer	Standard/Sample	Tracer	Antiserum
Blk	-	-	-	-
TA	-	-	5 μ l (at devl. step)	-
NSB	100 μ l	-	50 μ l	-
B ₀	50 μ l	-	50 μ l	50 μ l
Std/Sample	-	50 μ l	50 μ l	50 μ l

Table 1. Pipetting summary

Incubation of the Plate

Cover each plate with plastic film (Item No. 400012) and incubate overnight at 4°C. Optionally, the plate can be incubated for two hours at room temperature on an orbital shaker.

Development of the Plate

1. Reconstitute Ellman's Reagent immediately before use (20 ml of reagent is sufficient to develop 100 wells):

100 dtn vial Ellman's Reagent (96-well kit; Item No. 400050): Reconstitute with 20 ml of UltraPure water.

OR

250 dtn vial Ellman's Reagent (480-well kit; Item No. 400050): Reconstitute with 50 ml of UltraPure water.

NOTE: Reconstituted Ellman's Reagent is unstable and should be used the same day it is prepared; protect the Ellman's Reagent from light when not in use. Extra vials of the reagent have been provided should a plate need to be re-developed or multiple assays be run on different days.

2. Empty the wells and rinse five times with Wash Buffer.
3. Add 200 μl of Ellman's Reagent to each well
4. Add 5 μl of tracer to the TA wells.
5. Cover the plate with plastic film. Optimum development is obtained by using an orbital shaker equipped with a large, flat cover to allow the plate(s) to develop in the dark. This assay typically develops (i.e., B_0 wells ≥ 0.3 A.U. (blank subtracted)) in 90-120 minutes when using the 4°C overnight format. In the two hour room temperature format, the assay typically develops in 60-90 minutes.

Reading the Plate

1. Wipe the bottom of the plate with a clean tissue to remove fingerprints, dirt, etc.
2. Remove the plate cover being careful to keep Ellman's Reagent from splashing on the cover. *NOTE: Any loss of Ellman's Reagent will affect the absorbance readings. If Ellman's Reagent is present on the cover, use a pipette to transfer the Ellman's Reagent into the well. If too much Ellman's Reagent has splashed on the cover to easily redistribute back into the wells, wash the plate three times with wash buffer and repeat the development with fresh Ellman's Reagent.*
3. Read the plate at a wavelength between 405 and 420 nm. The absorbance may be checked periodically until the B_0 wells have reached a minimum of 0.3 A.U. (blank subtracted). The plate should be read when the absorbance of the B_0 wells in the range of 0.3-1.0 A.U. (blank subtracted). If the absorbance of the wells exceeds 1.5, wash the plate, add fresh Ellman's Reagent and let it develop again.

ANALYSIS

Many plate readers come with data reduction software that plots data automatically. Alternatively a spreadsheet program can be used. The data should be plotted as either $\%B/B_0$ versus log concentration using a four-parameter logistic fit or as logit B/B_0 versus log concentration using a linear fit. *NOTE: Cayman Chemical has a computer spreadsheet available for data analysis. Please contact Technical Service or visit our website (www.caymanchem.com/analysis/elisa) to obtain a free copy of this convenient data analysis tool.*

Calculations

Preparation of the Data

The following procedure is recommended for preparation of the data prior to graphical analysis.

NOTE: If the plate reader has not subtracted the absorbance readings of the blank wells from the absorbance readings of the rest of the plate, be sure to do that now.

1. Average the absorbance readings from the NSB wells.
2. Average the absorbance readings from the B_0 wells.
3. Subtract the NSB average from the B_0 average. This is the corrected B_0 or corrected maximum binding.
4. Calculate the B/B_0 (Sample or Standard Bound/Maximum Bound) for the remaining wells. To do this, subtract the average NSB absorbance from the S1 absorbance and divide by the corrected B_0 (from Step 3). Repeat for S2-S8 and all sample wells. (To obtain $\%B/B_0$ for a logistic four-parameter fit, multiply these values by 100.)

NOTE: The TA values are not used in the standard curve calculations. Rather, they are used as a diagnostic tool; the corrected B_0 divided by the actual TA (10X measured absorbance) will give the %Bound. This value should closely approximate the %Bound that can be calculated from the Sample Data (see pages 25 and 28). Erratic absorbance values and a low (or no) %Bound could indicate the presence of organic solvents in the buffer or other technical problems (see page 32 for Troubleshooting).

Plot the Standard Curve

Plot %B/B₀ for standards S1-S8 versus estriol concentration using linear (y) and log (x) axes and perform a 4-parameter logistic fit.

Alternative Plot - The data can also be linearized using a logit transformation. The equation for this conversion is shown below. *NOTE: Do not use %B/B₀ in this calculation.*

$$\text{logit (B/B}_0\text{)} = \ln [\text{B/B}_0\text{}/(1 - \text{B/B}_0\text{)}]$$

Plot the data as logit (B/B₀) versus log concentrations and perform a linear regression fit.

Determine the Sample Concentration

Calculate the B/B₀ (or %B/B₀) value for each sample. Determine the concentration of each sample by identifying the %B/B₀ on the standard curve and reading the corresponding values on the x-axis. *NOTE: Remember to account for any concentration of the sample prior to the addition to the well. Samples with %B/B₀ values greater than 80% or less than 20% should be re-assayed as they generally fall out of the linear range of the standard curve. A 20% or greater disparity between the apparent concentration of two different dilutions of the same sample indicates interference which could be eliminated by purification.*

Performance Characteristics

Sample Data

The standard curve presented here is an example of the data typically produced with this kit; however, your results will not be identical to these. You **must** run a new standard curve. Do not use the data below to determine the value of your samples. Your results could differ substantially.

	4°C Incubation (Overnight)			
	Raw Data		Average	Corrected
Total Activity	1.027	1.072	1.050	1.047
NSB	0.003	0.003	0.003	
B ₀	1.091	1.188		
	1.115	1.170	1.141	1.138

Dose (pg/ml)	Raw Data		Corrected		%B/B ₀	
2,000	0.086	0.083	0.083	0.080	7.3	7.0
666.7	0.151	0.149	0.148	0.146	12.9	12.8
222.2	0.263	0.283	0.260	0.280	22.8	24.6
74.1	0.425	0.450	0.422	0.447	37.1	39.3
24.7	0.629	0.629	0.626	0.626	55.0	55.0
8.2	0.832	0.802	0.829	0.799	72.8	70.2
2.7	0.964	0.960	0.961	0.957	84.5	84.1
0.9	1.066	1.053	1.063	1.050	93.4	92.3

Table 2. Typical results of overnight 4°C incubation




Figure 6. Typical standard curve of overnight 4°C incubation

Precision:

The intra- and inter-assay using an 18 hour incubation at 4°C. CVs have been determined at multiple points on the standard curve. These data are summarized in the graph on page 26 and in the table below.

Dose (pg/ml)	%CV* Intra-assay variation	%CV* Inter-assay variation
2,000	15.0	13.9
666.7	9.5	6.8
222.2	11.2	6.5
74.1	11.9	9.2
24.7	11.1	15.6
8.2	17.9	25.7
2.7	†	†
0.9	†	†

Table 3. Intra- and inter-assay variation of overnight 4°C incubation

*%CV represents the variation in concentration (not absorbance) of each point on the standard curve as determined using a reference standard curve.

†Outside of the recommended usable range of the assay.


Room Temperature Incubation (Two hours)

	Raw Data		Average	Corrected
Total Activity	0.723	0.785	0.754	0.744
NSB	0.010	0.009	0.010	
B_0	0.421	0.408		
	0.426	0.439	0.424	0.414

Dose (pg/ml)	Raw Data		Corrected		%B/ B_0	
2,000	0.069	0.069	0.059	0.060	14.3	14.5
666.7	0.109	0.102	0.099	0.093	22.4	22.4
222.2	0.164	0.173	0.155	0.164	39.5	39.5
74.1	0.230	0.230	0.221	0.221	53.3	53.3
24.7	0.310	0.310	0.300	0.301	72.6	72.6
8.2	0.362	0.354	0.353	0.345	83.3	83.3
2.7	0.401	0.403	0.392	0.394	95.1	95.1
0.9	0.437	0.427	0.427	0.423	102.3	102.3

Table 4. Typical results of two hour room temperature incubation

Estriol Two Hours at Room Temperature



Assay Range = 0.9-2,000 pg/ml
Sensitivity (defined as 80% B/ B_0) = 12 pg/ml
 The sensitivity and mid-point were derived from the standard curve shown above. The standard was diluted with ELISA Buffer.

Figure 7. Typical standard curve of two hour room temperature incubation

Precision:

The intra- and inter-assay using a two hour incubation at room temperature. CVs have been determined at multiple points on the standard curve. These data are summarized in the graph on page 29 and in the table below.

Dose (pg/ml)	%CV*	
	Intra-assay variation	Inter-assay variation
2,000	20.9	9.2
666.7	16.1	7.6
222.2	11.3	10.4
74.1	11.6	11.9
24.7	15.9	16.9
8.2	17.0	16.8
2.7	†	†
0.9	†	†

Table 5. Intra- and inter-assay variation of two hour room temperature incubation

*%CV represents the variation in concentration (not absorbance) of each point on the standard curve as determined using a reference standard curve.

†Outside of the recommended usable range of the assay.

Cross Reactivity:

Compound	Cross Reactivity
Estriol	100%
Estriol-3-glucuronide	6%
Estriol-16-glucuronide	1.2%
Estradiol	0.6%
Estrone	0.02%
11-deoxy Corticosterone	<0.01%
Cortisol	<0.01%
DHEA	<0.01%
Ethinyl estradiol	<0.01%
Progesterone	<0.01%
Testosterone	<0.01%
5 α -dihydro Testosterone	<0.01%

Table 6. Cross Reactivity of the Estriol ELISA

RESOURCES

Troubleshooting

Problem	Possible Causes	Recommended Solutions
Erratic values; dispersion of duplicates	A. Trace organic contaminants in the water source B. Poor pipetting/technique	A. Replace activated carbon filter or change source of UltraPure water
High NSB (>10% of B ₀)	A. Poor washing B. Exposure of NSB wells to specific antibody	A. Re-wash plate and re-develop
Very low B ₀	A. Trace organic contaminants in the water source B. Plate requires additional development time C. Dilution error in preparing reagents	A. Replace activated carbon filter or change source of UltraPure water B. Return plate to shaker and re-read later
Low sensitivity (shift in dose response curve)	Standard is degraded	Replace standard
Analyses of two dilutions of a biological sample do not agree (i.e., more than 20% difference)	Interfering substances are present	Purify sample prior to analysis by ELISA ⁶
Only Total Activity (TA) wells develop	Trace organic contaminants in the water source	Replace activated carbon filter or change source of UltraPure water

References

1. Peter, M., Dörr, H.G., and Sippell, W.G. Changes in the concentrations of dehydroepiandrosterone sulfate and estriol in maternal plasma during pregnancy: A longitudinal study in healthy women throughout gestation and at term. *Horm. Res.* **42**, 278-281 (1994).
2. Goodwin, T.M. A role for estriol in human labor, term and preterm. *Amer. J. Obstet. Gynecol.* **180**, S208-S213 (1999).
3. Voss, H.F. Hormonal pathways of preterm birth. *Amer. J. Obstet. Gynecol.* **180**, S226-S231 (1999).
4. Maclouf, J., Grassi, J., and Pradelles, P. Development of enzyme-immunoassay techniques for the measurement of eicosanoids, Chapter 5, in *Prostaglandin and Lipid Metabolism in Radiation Injury*. Walden, T.L., Jr. and Hughes, H.N., editors, Plenum Press, Rockville, 355-364 (1987).
5. Pradelles, P., Grassi, J. and Maclouf, J. Enzyme immunoassays of eicosanoids using acetylcholinesterase as label: An alternative to radioimmunoassay. *Anal. Chem.* **57**, 1170-1173 (1985).
6. Maxey, K.M., Maddipati, K.R. and Birkmeier, J. Interference in enzyme immunoassays. *J. Clin. Immunoassay* **15**, 116-120 (1992).

1								
2								
3								
4								
5								
6								
7								
8								
9								
10								
11								
12								
	A	B	C	D	E	F	G	H

NOTES

Warranty and Limitation of Remedy

Buyer agrees to purchase the material subject to Cayman's Terms and Conditions. Complete Terms and Conditions including Warranty and Limitation of Liability information can be found on our website.

This document is copyrighted. All rights are reserved. This document may not, in whole or part, be copied, photocopied, reproduced, translated, or reduced to any electronic medium or machine-readable form without prior consent, in writing, from Cayman Chemical Company.

©08/15/2016, Cayman Chemical Company, Ann Arbor, MI, All rights reserved.
Printed in U.S.A.

