

**FLUOROSHIELD™ Mounting Medium**

**Catalog number:** AR-6500-01 (20 ml), AR-6500-02 (100 ml), AR-6500-03 (250 ml), AR-6500-05 (1000 ml).

**Description:** Fluoroshield is an aqueous mounting medium for preserving fluorescence of tissue and cell smears. This unique formula prevents rapid photobleaching of FITC, Texas Red, AMCA, Cy2, Cy3, Cy5, Alexa fluoro 488, Alexa fluoro 594, Green fluorescent protein (GFP), tetramethyl rhodamine, Redox, Phycoerythrin (R-PE), phycocyanin (PC), and allophycocyanin (APC). The fluorescence is retained during prolonged storage at 4°C in the dark. This medium does not contain phenylenediamine, which destroys immunofluorescence of Cy dyes, R-PE, PC and APC.

**Intended Use:** Immunofluorescence, confocal microscopy.

**Reagent:** Ready to use mounting medium

**Storage:** This product can be store at room temp, however for long term storage 2-8°C is recommended. Protect from light, DO NOT FREEZE.

**Procedure:**

1. Bring the vial to room temperature.
2. Rinse slide to be mounted with DISTILLED OR DEIONIZED WATER, touch the edges of slide on a paper towel to remove excess water. Place slides on a flat surface away from light.
3. Turn the vial upside down and open the dropper to remove any air bubbles.
4. Apply 3-4 drops of mounting medium directly on top of the specimen and spread out evenly by tilting slide back and forth or spread evenly with a 0.2 ml plastic pipette tip making sure the tissue is not touched. Excess medium can be removed by touching the edges of slide against paper towel.
5. Let stand at room temperature for about 5 minutes.
6. Apply cover slip carefully avoiding air bubbles.
7. The specimen is ready for visualization under a microscope.
8. One can seal the edges of cover slip with nail polish, any organic medium or our Organo mounting medium (cat. # AR-6504). If a coverslip is not used air bubbles will appear in few days.
9. For storage it is recommended that the slide be stored in the dark at 2-8°C.

**Removal of Coverslip:** *Coverslip can be removed before sealing the edges. Soak slide in warm (37°C) water for several minutes. Carefully and slowly move the coverslip. Soak in water for an additional few minutes to remove coverslip. Rinse slide several times with warm water to remove all mounting medium. The slide can be remounted again.*

**Limitation and warranty:** Our warranty is limited to the actual price paid for the product. We are not liable for any property damage, personnel injury, time, effort or economic loss due to our product.

**MSDS:** This product contains 0.05 % sodium azide as a preservative, appropriate care should be taken in handling. National Institute of Occupational Safety and Health has warning that sodium azide can react with lead, copper, brass or solder in the plumbing system and forms hydrazoic acid in acidic condition. Discard with copious amount of water. Avoid skin and eye contact with all laboratory products. Use appropriate lab. gear, lab coat, gloves and safety glasses. Do not ingest any lab. products. This product is not approved for administration in human or animals.

---

Fluoroshield is a trademark of ImmunoBioScience corp.

For research use only, not for use in diagnostic or therapeutic procedures or in vivo use

Phone: 425 367 4601; 425 514 3761; Fax: 425 367 4817; cell (mobile) phone: 425 314 0199

Web site: [www.immunobioscience.com](http://www.immunobioscience.com)

E-mail: [baderbo@gmail.com](mailto:baderbo@gmail.com) or [anitaIBSC@aol.com](mailto:anitaIBSC@aol.com)