


IDK[®] anti-SARS-CoV-2 IgG ELISA

*For the qualitative determination of
human IgG antibodies against SARS-CoV-2
in EDTA plasma and serum*

Valid from 2021-04-16

REF KR5000

Σ 96



RUO



Immundiagnostik AG, Stubenwald-Allee 8a, 64625 Bensheim, Germany

Tel.: +49 6251 70190-0

Fax: + 49 6251 70190-363

e.mail: info@immundiagnostik.com

www.immundiagnostik.com

Table of Contents

1. INTENDED USE	2
2. INTRODUCTION	2
3. MATERIAL SUPPLIED	3
4. MATERIAL REQUIRED BUT NOT SUPPLIED	3
5. PREPARATION AND STORAGE OF REAGENTS	3
6. PREPARATION OF THE ASSAY	4
<i>Sample storage</i>	4
<i>Dilution of samples</i>	4
7. ASSAY PROCEDURE	4
<i>Principle of the test</i>	4
<i>Test procedure</i>	5
8. RESULTS	6
9. LIMITATIONS	6
10. QUALITY CONTROL	6
11. PERFORMANCE CHARACTERISTICS	7
<i>Clinical specificity and sensitivity</i>	7
<i>Accuracy – Precision</i>	7
12. PRECAUTIONS	7
13. TECHNICAL HINTS	8
14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE	8
15. LITERATURE	9

1. INTENDED USE

IDK® anti-SARS-CoV-2 IgG is an enzyme-linked immunosorbent assay (ELISA) for qualitative measurement of IgG antibodies against the novel coronavirus SARS-CoV-2 in human serum or EDTA plasma. The assay is a research device and is intended to be used by professional users in a laboratory environment. This ELISA can be performed manually or using an automated platform. This test serves as complement to infection diagnosis and provides evidence for a previous infection with SARS-CoV-2.

2. INTRODUCTION

The virus SARS-CoV-2 (severe acute respiratory syndrome coronavirus 2) is an enveloped, single stranded RNA virus and is a family member of the coronavirus family *Coronaviridae* [1,2].

Coronaviruses have a similar composition as they are composed of similar structural proteins including the spike (S), envelope (E), membrane (M), and nucleocapsid (N) protein and several non-structural proteins. Their members cause a variety of diseases in different vertebrate species [6, 8]. As of February 2020, seven human pathogenic coronaviruses are known: besides SARS-CoV[-1], SARS-CoV-2 and MERS-CoV, there are HCoV-HKU1, HCoV-OC43, HCoV-NL63 and HCoV-229E. While SARS-CoV[-1], SARS-CoV-2 and MERS-CoV can cause severe respiratory and systemic diseases, infections with the last four mentioned usually only lead to mild cold symptoms [3, 8].

After an infection with SARS-CoV-2, the virus accesses host cells via the protein ACE2 (angiotensin-converting enzyme) and causes the disease COVID-19. The severity of disease ranges from asymptomatic, mild (fever, cold, cough, tiredness, shortness of breath, and loss of smell), and severe to most severe forms with death [4, 5, 7]. Aging and several co-morbidities (e.g. diabetes mellitus, cardiovascular diseases, and chronic pulmonary diseases) are described as risk factors for severe progressive forms of COVID-19 [3, 12, 13].

Seroconversion occurs on different time points depending on the used method and the measured class of antibodies. The onset of IgG antibodies is usually observed after 11 to 14 days and seroconversion of IgG antibodies rises its maximum after three to six weeks [9, 10, 11].

3. MATERIAL SUPPLIED

Cat. No.	Label	Kit components	Quantity
KR5 000	PLATE	Microtiter plate, pre-coated	12 x 8 wells
KR0001.C.100	WASHBUF	Wash buffer concentrate, 10 x	2 x 100 ml
KR5000	CONJ	Conjugate, ready-to-use	1 x 12.5 ml
KR5000	CTRL CUT-OFF	Cut-off control, ready-to-use	1 x 1 ml
KR5000	CTRL NEG	Negative control, ready-to-use	1 x 1 ml
KR5000	CTRL POS	Positive control, ready-to-use	1 x 1 ml
KR5000	SAMPLEBUF	Sample dilution buffer, ready-to-use	1 x 110 ml
KR0002.15	SUB	Substrate (tetramethylbenzidine), ready-to-use	1 x 15 ml
KR0003.15	STOP	Stop solution, ready-to-use	1 x 15 ml
	FOL	Foil to cover the microtiter plate	3 x

For reorders of single components, use the catalogue number followed by the label as product number.

4. MATERIAL REQUIRED BUT NOT SUPPLIED

- Ultrapure water*
- Calibrated precision pipettors and 10–1 000 µl single-use tips
- Horizontal microtiter plate shaker
- A multi-channel dispenser or repeating dispenser
- Vortex
- Standard single-use laboratory glass or plastic vials, cups, etc.
- Microtiter plate reader (required filters see chapter 7)

* Immundiagnostik AG recommends the use of ultrapure water (water type 1; ISO 3696), which is free of undissolved and colloidal ions and organic molecules (free of particles > 0.2 µm) with an electrical conductivity of 0.055 µS/cm at 25 °C (≥ 18.2 MΩ cm).

5. PREPARATION AND STORAGE OF REAGENTS

- **Bring all reagents to room temperature (18–30 °C) prior to use.**
- To run the assay more than once, ensure that reagents are stored at conditions stated on the label. **Prepare only the appropriate amount necessary for each run.** The kit can be used up to 4 times within the expiry date stated on the label.

- **Preparation of the wash buffer:** The **wash buffer concentrate (WASHBUF)** should be diluted with ultrapure water **1:10** before use (100 ml WASHBUF + 900 ml ultrapure water), mix well. Crystals could occur due to high salt concentration in the concentrate. The crystals must be redissolved at room temperature or in a water bath at 37°C before dilution of the buffer solutions. The **WASHBUF** is stable at **2–8 °C** until the expiry date stated on the label. **Wash buffer** (1:10 diluted WASHBUF) can be stored in a closed flask at **2–8 °C for 1 month**.
- All other test reagents are ready to use. Test reagents are stable until the expiry date (see label) when stored at **2–8 °C**.

6. PREPARATION OF THE ASSAY

Sample storage

Freshly collected serum can be stored for 14 days at room temperature or for up to 4 weeks at 2–8°C. Long-term storage is recommended at -20°C. More than 3 freeze-thaw cycles should be avoided.

Diluted samples are not stable and cannot be stored.

Dilution of samples

Samples are diluted **1:101 in sample dilution buffer**. For example:

- **10 µl** sample + **1 000 µl** sample dilution buffer, mix well = **1:101**

For analysis, pipet **100 µl diluted sample** per well.

7. ASSAY PROCEDURE

Principle of the test

This ELISA serves for the determination of IgG antibodies against the spike protein (S1) of SARS-CoV-2. Diluted samples are added to a microtiter plate coated with a specific antigen. By adding the peroxidase conjugate (peroxidase labelled detection antibody), the antibodies against SARS-CoV-2 in the sample are marked. They are detected via the peroxidase conjugate with the peroxidase converting the substrate TMB to a blue product. The enzymatic reaction is stopped by adding an acidic solution. The samples convert from blue to yellow. The colour change should be measured in a photometer at 450 nm. The interpretation is made using the cut-off control.

Test procedure

Bring all reagents and samples to room temperature (18–30 °C) and mix well.

Mark the positions of controls/samples on a protocol sheet.

Take as many microtiter strips (PLATE) as needed from kit. Store unused strips covered with the foil included in the kit together with the desiccant bag in the re-closed aluminium packaging at 2–8 °C. Strips are stable until the expiry date stated on the label.

For automated ELISA processors, the given protocol may need to be adjusted according to the specific features of the respective automated platform. For further details, please contact your supplier or Immundiagnostik AG.

We recommend carrying out the tests in duplicates.

1.	Pipet each 100 µl of controls and diluted samples into the wells of the microtiter plate.
2.	Cover the strips and incubate for 1 hour shaking (900 rpm)* on a horizontal shaker at room temperature (18–30 °C).
3.	Discard the content of each well and wash 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
4.	Add 100 µl conjugate (CONJ) into each well.
5.	Cover the strips and incubate for 1 hour shaking (900 rpm)* on a horizontal shaker at room temperature (18–30 °C).
6.	Discard the content of each well and wash 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
7.	Add 100 µl substrate (SUB) into each well.
8.	Incubate for 10–15 minutes at room temperature (18–30 °C) until a sufficiently large difference in colour occurs**.
9.	Add 100 µl stop solution (STOP) into each well and mix shortly by using the shake function (900 rpm)* of the microplate reader.
10.	Determine absorption immediately with an ELISA reader at 450 nm against 620 nm (or 690 nm) as a reference.

* We recommend shaking the strips with an orbit of 2 mm.

** The intensity of the colour change is temperature sensitive. We recommend observing the colour change and stopping the reaction upon good differentiation.

8. RESULTS

Cut-off = OD cut-off control x 1.2

Samples which have a higher average optical density (OD) than the **cut-off** are positive.

Samples which have a lower average optical density (OD) than the **cut-off** are negative.

The plausibility of the pairs of values should be examined before the automatic evaluation of the results. If this option is not available with the used program, a manual control of the paired values should be made.

9. LIMITATIONS

Samples which cannot be clearly interpreted (e.g. because of high coefficients of variation of replicates) should be measured again.

Negative IgG results do not rule out an infection with SARS-CoV-2. The serum or plasma samples may be collected at a very early stage of infection when the body has not yet produced IgG antibodies. These are produced about 11–14 days after the start of an infection. They arise at late infection stages or after an infection has been overcome. Therefore, this test cannot be used to diagnose an acute infection.

In case of only weak positive results, it is recommended to re-sample and analyse after ~14 days.

10. QUALITY CONTROL

Immundiagnostik AG recommends the use of external controls for internal quality control, if possible.

Control samples should be analysed with each run. Results, generated from the analysis of control samples, should be evaluated for acceptability using appropriate statistical methods. The results for the samples may not be valid if within the same assay one or more values of the quality control sample are outside the acceptable limits.

11. PERFORMANCE CHARACTERISTICS

Clinical specificity and sensitivity

Accuracy – Precision

Repeatability (Intra-Assay)

The repeatability was assessed with 1 serum sample and the supplied cut-off control under **constant** parameters (same operator, instrument, day and kit lot).

Sample	Mean value [OD]	CV [%]
1 (n = 74)	1.943	2.92 %
2 (CTRL CUT-OFF) (n = 80)	0.396	3.03 %

Reproducibility (Inter-Assay)

The reproducibility was assessed with 1 serum sample under **varying** parameters (different operators, instruments, days and kit lots).

Sample	Mean value [OD]	CV [%]
1 (n = 12)	1.457	5.5
2 (n = 13)	0.121	8.4

12. PRECAUTIONS

- All reagents in the kit package are for research use only.
- Human materials used in kit components were tested and found to be negative for HIV, Hepatitis B and Hepatitis C. However, for safety reasons, all kit components should be treated as potentially infectious.
- Kit reagents contain sodium azide or ProClin as bactericides. Sodium azide and ProClin are toxic. Substrates for the enzymatic colour reactions are toxic and carcinogenic. Avoid contact with skin or mucous membranes.
- The stop solution consists of diluted sulphuric acid, a strong acid. Although diluted, it still must be handled with care. It can cause burns and should be handled with gloves, eye protection, and appropriate protective clothing. Any

spill should be wiped up immediately with copious quantities of water. Do not breath vapour and avoid inhalation.

13. TECHNICAL HINTS

- Do not interchange different lot numbers of any kit component within the same assay. Furthermore, we recommend not assembling wells of different microtiter plates for analysis, even if they are of the same batch.
- Control samples should be analysed with each run.
- Reagents should not be used beyond the expiration date stated on kit label.
- Substrate solution should remain colourless until use.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- Avoid foaming when mixing reagents.
- Do not mix plugs and caps from different reagents.
- The assay should always be performed according to the enclosed manual.

14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE












- The guidelines for laboratories should be followed.
- *IDK®* is a trademark of Immundiagnostik AG.
- Incubation time, incubation temperature and pipetting volumes of the components are defined by the producer. Any variation of the test procedure, which is not coordinated with the producer, may influence the results of the test. Immundiagnostik AG can therefore not be held responsible for any damage resulting from incorrect use.
- Warranty claims and complaints regarding deficiencies must be logged within 14 days after receipt of the product. The product should be send to Immundiagnostik AG along with a written complaint.

15. LITERATURE

1. Coronaviridae Study Group of the International Committee on Taxonomy of Viruses. 2020. "The Species Severe Acute Respiratory Syndrome-Related Coronavirus: Classifying 2019-NCoV and Naming It SARS-CoV-2." *Nature Microbiology* **5** (4): 536–44.
2. Zhu, Na, Dingyu Zhang, Wenling Wang, Xingwang Li, Bo Yang, Jingdong Song, Xiang Zhao, et al. 2020. "A Novel Coronavirus from Patients with Pneumonia in China, 2019." *The New England Journal of Medicine* **382** (8): 727–33.
3. Tian, Wenjie, Wanlin Jiang, Jie Yao, Christopher J Nicholson, Rebecca H Li, Haakon H Sigursslid, Luke Wooster, Jerome I Rotter, Xiuqing Guo, and Rajeev Malhotra. 2020. "Predictors of Mortality in Hospitalized COVID-19 Patients: A Systematic Review and Meta-Analysis." *Journal of Medical Virology*, May.
4. Bourgonje, Arno R, Amaal E Abdulle, Wim Timens, Jan-Luuk Hillebrands, Gerjan J Navis, Sanne J Gordijn, Marieke C Bolling, et al. 2020. "Angiotensin-Converting Enzyme 2 (ACE2), SARS-CoV-2 and the Pathophysiology of Coronavirus Disease 2019 (COVID-19)." *The Journal of Pathology*, May.
5. Yan, Renhong, Yuanyuan Zhang, Yaning Li, Lu Xia, Yingying Guo, and Qiang Zhou. 2020. "Structural Basis for the Recognition of SARS-CoV-2 by Full-Length Human ACE2." *Science (New York, N.Y.)* **367** (6485): 1444–48.
6. Wu, Aiping, Yousong Peng, Baoying Huang, Xiao Ding, Xianyue Wang, Peihua Niu, Jing Meng, et al. 2020. "Genome Composition and Divergence of the Novel Coronavirus (2019-NCoV) Originating in China." *Cell Host & Microbe* **27** (3): 325–28.
7. Hoffmann, Markus, Hannah Kleine-Weber, Simon Schroeder, Nadine Krüger, Tanja Herrler, Sandra Erichsen, Tobias S Schiergens, et al. 2020. "SARS-CoV-2 Cell Entry Depends on ACE2 and TMPRSS2 and Is Blocked by a Clinically Proven Protease Inhibitor." *Cell* **181** (2): 271-280.e8.
8. Corman, V M, J Lienau, and M Witzenth. 2019. "[Coronaviruses as the Cause of Respiratory Infections]." *Der Internist* **60** (11): 1136–45.
9. Wölfel, Roman, Victor M Corman, Wolfgang Guggemos, Michael Seilmaier, Sabine Zange, Marcel A Müller, Daniela Niemeyer, et al. 2020. "Virological Assessment of Hospitalized Patients with COVID-2019." *Nature* **581** (7809): 465–69.
10. To, Kelvin Kai-Wang, Owen Tak-Yin Tsang, Wai-Shing Leung, Anthony Raymond Tam, Tak-Chiu Wu, David Christopher Lung, Cyril Chik-Yan Yip, et al. 2020. "Temporal Profiles of Viral Load in Posterior Oropharyngeal Saliva Samples and Serum Antibody Responses during Infection by SARS-CoV-2: An Observational Cohort Study." *The Lancet. Infectious Diseases* **20** (5): 565–74.

11. Xiao, Ai Tang, Chun Gao, and Sheng Zhang. 2020. "Profile of Specific Antibodies to SARS-CoV-2: The First Report." *The Journal of Infection* **81** (1): 147–78.
12. Sheervalilou, Roghayeh, Milad Shirvaliloo, Nahid Dadashzadeh, Sakine Shirvalilou, Omolbanin Shahraki, Younes Pilehvar-Soltanahmadi, Habib Ghaznavi, Samideh Khoei, and Ziba Nazarlou. 2020. "COVID-19 under Spotlight: A Close Look at the Origin, Transmission, Diagnosis, and Treatment of the 2019-NCoV Disease." *Journal of Cellular Physiology*, May.
13. Lotfi, Melika, and Nima Rezaei. 2020. "SARS-CoV-2: A Comprehensive Review from Pathogenicity of the Virus to Clinical Consequences." *Journal of Medical Virology*, June.
14. Eberhardt, K.A.; Dewald, F.; Heger, E.; Gieselmann, L.; Vanshylla, K.; Wirtz, M.; Kleipass, F.; Johannis, W.; Schommers, P.; Gruell, H.; Brensing, K.A.; Müller, R.-U.; Augustin, M.; Lehmann, C.; Koch, M.; Klein, F.; Di Cristanziano, V. Evaluation of a New Spike (S)-Protein-Based Commercial Immunoassay for the Detection of Anti-SARS-CoV-2 IgG. *Microorganisms* 2021, 9, 733.

Used symbols:

	Temperature limitation		Catalogue Number
	For research use only		To be used with
	Manufacturer		Contains sufficient for <n> tests
	Lot number		Use by
	Attention		Consult instructions for use
	Consult specification data sheet		