



CRP ELISA

*For the determination of C-reactive protein
in serum, plasma, dried blood spots, stool and urine*

Valid from 2019-07-18



KR9720s



Immundiagnostik AG, Stubenwald-Allee 8a, 64625 Bensheim, Germany

Tel.: +49 6251 70190-0

Fax: + 49 6251 70190-363

e.mail: info@immundiagnostik.com

www.immundiagnostik.com

Table of Contents

1. INTENDED USE	2
2. INTRODUCTION	2
3. MATERIAL SUPPLIED	2
4. MATERIAL REQUIRED BUT NOT SUPPLIED	2
5. STORAGE AND PREPARATION OF REAGENTS	3
6. STORAGE AND PREPARATION OF SAMPLES	4
<i>Serum and plasma</i>	4
<i>Collection and storage of samples</i>	4
<i>Dried blood spots</i>	4
<i>Urine samples</i>	5
<i>Stool samples</i>	5
7. ASSAY PROCEDURE	6
<i>Principle of the test</i>	6
<i>Test procedure</i>	7
8. RESULTS	8
9. LIMITATIONS	9
10. QUALITY CONTROL	9
<i>Reference range</i>	9
11. PERFORMANCE CHARACTERISTICS	10
<i>Analytical Sensitivity</i>	10
<i>Precision and reproducibility</i>	10
12. PRECAUTIONS	10
13. TECHNICAL HINTS	11
14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE	11
15. REFERENCES	11

1. INTENDED USE

This Immundiagnostik AG assay is an enzyme immunoassay intended for the quantitative determination of C-reactive Protein in plasma, serum, dried blood spots, stool and urine. For research use only. Not for use in diagnostic procedures.

2. INTRODUCTION

C-reactive Protein (CRP) is mainly formed in hepatocytes. The synthesis rate of CRP is influenced by the cytokines involved in the inflammatory processes. The biological half-life time is estimated to be 13–16 hours.

3. MATERIAL SUPPLIED

Cat. No.	Label	Kit components	Quantity
KR9720s	PLATE	Microtiter plate, pre-coated	12 x 8 wells
KR0001.C.100	WASHBUF	Wash buffer concentrate, 10x	1 x 100 ml
KR9720s	CONJ	Conjugate, (rabbit-anti-CRP-antibody, peroxidase-labelled)	1 x 150 µl
KR9720s	CAL	Calibrator*, ready-to-use	1 x 1 ml
KR9720s	CTRL 1	Control, ready-to-use (see specification for range)	1 x 1 ml
KR9720s	CTRL 2	Control, ready-to-use (see specification for range)	1 x 1 ml
KR9720s	SAMPLEBUF	Sample dilution buffer, ready-to-use	2 x 100 ml
KR0002.15	SUB	Substrate (tetramethylbenzidine), ready-to-use	1 x 15 ml
KR0003.15	STOP	Stop solution, ready-to-use	1 x 15 ml

For reorders of single components, use the catalogue number followed by the label as product number.

*The CRP calibrators were standardised against WHO standard 470.

4. MATERIAL REQUIRED BUT NOT SUPPLIED

- Ultrapure water*
- Dried blood spot carrier such as DrySpot-ID cat. no.: DZ9020ID or DZ9021ID
- Stool sample application system such as Cat. No.: KR6998SAS
- Calibrated precision pipettors and 10–1000 µl single-use tips

- Foil to cover the microtiter plate
- Horizontal microtiter plate shaker
- Multi-channel pipets or repeater pipets
- Centrifuge
- Vortex
- Standard single-use laboratory glass or plastic vials, cups, etc.
- Microtiter plate reader (required filters see chapter 7)

* Immundiagnostik AG recommends the use of Ultra Pure Water (Water Type 1; ISO 3696), which is free of undissolved and colloidal ions and organic molecules (free of particles > 0.2 µm) with an electrical conductivity of 0.055 µS/cm at 25 °C (≥ 18.2 MΩ cm).

5. STORAGE AND PREPARATION OF REAGENTS

- To run the assay more than once, ensure that reagents are stored at the conditions stated on the label. **Prepare only the appropriate amount necessary for each run.** The kit can be used up to 4 times within the expiry date stated on the label.
- Reagents with a volume less than **100 µl** should be centrifuged before use to avoid loss of volume.
- **Preparation of the wash buffer:** The **wash buffer concentrate (WASHBUF)** has to be diluted with ultrapure water **1:10** before use (100 ml WASHBUF + 900 ml ultrapure water), mix well. Crystals could occur due to high salt concentration in the concentrate. Before dilution, the crystals have to be redissolved at room temperature or in a water bath at 37 °C. The **WASHBUF** is stable at **2–8 °C** until the expiry date stated on the label. **Wash buffer** (1:10 diluted WASHBUF) can be stored in a closed flask at **2–8 °C for 1 month**.
- Use **100 µl** of **wash buffer** (1:10 diluted WASHBUF) as **blank**.
- **Preparation of the conjugate:** Before use, the **conjugate concentrate (CONJ)** has to be diluted **1:101** in wash buffer (100 µl CONJ + 10 ml wash buffer). The CONJ is stable at **2–8 °C** until the expiry date stated on the label. **Conjugate** (1:101 diluted CONJ) **is not stable and cannot be stored**.
- All other test reagents are ready-to-use. Test reagents are stable until the expiry date (see label) when stored at **2–8 °C**.

6. STORAGE AND PREPARATION OF SAMPLES

Serum and plasma

Collection and storage of samples

Collection of serum

Collect sufficient blood (at least 1 ml) by venipuncture into a tube or a plastic syringe, avoid hemolysis, allow to stand for 15 min, centrifuge for 15 min at 1,000 g and 4 °C and collect the serum.

Collection of plasma

Collect sufficient blood (at least 1 ml) by venipuncture into an EDTA venipuncture tube or a plastic syringe, allow to stand for 15 min, centrifuge for 15 min at 1,000 g and 4 °C and separate the plasma from the cells.

Storage of serum

Serum samples can be stored at -80 °C for 11 years.*

*A. P. Doumatey et al. 2014.

Sample dilution

Serum and plasma samples have to be diluted **1:100 or 1:500** before performing the assay.

For a dilution of 1:100 e.g.:

Add **10 µl serum /plasma to 990 µl sample dilution buffer** (SAMPLEBUF) and mix well.

Samples with **elevated CRP-concentrations** must be diluted **1:4 000–1:8 000**. Samples of other collectives must be diluted according to the expected CRP-concentration.

Dried blood spots

Collection and storage of dried blood spots

We recommend DrySpot-ID (catalogue no DZ9020ID or DZ9021ID) as dried blood spot carrier. The moistened cards are stable for 2 weeks at room temperature.

Preparation of dried blood samples

1.	Label 1,5- ml polypropylene tubes.
2.	Remove filter from sampling device.
3.	Put filter in a labelled tube.

4.	Add 400 µl sample dilution buffer (SAMPLEBUF) to each sample, allow sample to stand for 20 min at room temperature (15–30°C).
5.	Vortex for 10 s . The filter will decolourise.

For testing in duplicates, pipette 2 x 100 µl of each prepared sample per well.

Urine samples

Storage of urine samples

Urine should be stored at –20°C until the measurement. CRP in urine is stable for 4 weeks at –20°C.

Dilution of urine samples

Urine samples must be diluted **1:5** before performing the assay,

e.g. **50 µl** sample + **200 µl** sample dilution buffer (SAMPLEBUF), mix well.

For analysis, pipet 100 µl of this dilution per well.

Stool samples

Storage of stool

The samples should be refrigerated and can be stored at **2-8°C for 2 days**. If the test cannot be performed within this period, the specimen should be stored at –20°C or colder.

Extraction of the stool samples

Wash buffer (diluted WASHBUF) is used as a sample extraction buffer. We recommend the following sample preparation:

Stool Sample Application System (SAS) (Cat. No.: KR6998SAS)

Stool sample tube – Instructions for use

Please note that the dilution factor of the final stool suspension depends on the amount of stool sample used and the volume of the buffer.

SAS with 0.75 ml wash buffer:

Applied amount of stool:	15 mg
Buffer Volume:	0.75 ml
Dilution Factor:	1:50

Please follow the instructions for the preparation of stool samples using the SAS as follows:

- a) The raw stool sample has to be thawed. For particularly heterogeneous samples we recommend a mechanical homogenisation using an applicator, inoculation loop or similar device.
- b) Fill the **empty stool sample tube** with **0.75 ml wash buffer** (diluted WASH-BUF) before using it with the sample. **Important:** Allow the extraction buffer to reach room temperature.
- c) Unscrew the tube (yellow part of cap) to open. Insert the yellow dipstick into the sample. The lower part of the dipstick has notches which need to be covered completely with stool after inserting it into the sample. Place dipstick back into the tube. When putting the stick back into the tube, excess material will be stripped off, leaving 15 mg of sample to be diluted. Screw tightly to close the tube.
- d) Shake the tube well until no stool sample remains in the notches. Important: Please make sure that you have a maximally homogenous suspension after shaking. Especially with more solid samples, soaking the sample in the tube with buffer for ~ 10 minutes improves the result.
- e) Allow sample to stand for ~10 minutes until sediment has settled. Floating material like shells of grains can be neglected.
- f) Carefully unscrew the complete cap of the tube including the blue ring plus the dipstick. Discard cap and dipstick. Make sure that the sediment will not be dispersed again.

Dilution: 1:50

The extract can be stored **1 month at -20°C**.

100 µl per well of this supernatant are used in the assay.

7. ASSAY PROCEDURE

Principle of the test

This ELISA is designed for the quantitative determination of CRP in serum, plasma, urine and stool samples. The wells of the microtiter plate are coated with polyclonal antibodies directed against C-reactive Protein. In a first incubation step, the CRP in the samples is bound to the coated polyclonal rabbit antibodies (in excess). To remove all unbound substances, a washing step is carried out.

In a second incubation step, a peroxidase-labelled antibody (polyclonal, rabbit-anti-CRP) is added. After another washing step, to remove all unbound substances, the solid phase is incubated with the substrate, tetramethylbenzidine. An acidic stopping solution is then added. The colour converts to yellow. The intensity of the yellow colour is directly proportional to the concentration of CRP in the sample. A dose response curve of the absorbance unit (optical density, OD at 450 nm) vs. concentration is generated, using the values obtained from the standard. CRP, present in the samples, is determined directly from this curve.

The combination of two specific antibodies in the CRP ELISA drastically reduces the possibility of wrong-negatives results.

Test procedure

Bring all **reagents and samples to room temperature** (15–30 °C) and mix well.

Mark the positions of calibrator/controls/blank/samples on a protocol sheet.

Take as many microtiter strips as needed from the kit. Store unused strips together with the desiccant bag in the closed aluminium packaging at 2–8 °C. Strips are stable until expiry date stated on the label.

For automated ELISA processors, the given protocol may need to be adjusted according to the specific features of the respective automated platform. For further details please contact your supplier or Immundiagnostik AG.

We recommend to carry out the tests in duplicate.

1.	Before use , wash the wells 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
2.	Add each 100 µl calibrator/controls/blank/prepared samples into the respective wells.
3.	Cover the strips and incubate for 1 hour at room temperature (15–30 °C) on a horizontal shaker** .
4.	Discard the content of each well and wash 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
5.	Add 100 µl conjugate (diluted CONJ) into each well.
6.	Cover the strips and incubate for 1 hour at room temperature (15–30 °C) on a horizontal shaker** .

7.	Discard the content of each well and wash 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
8.	Add 100 µl substrate (SUB) into each well.
9.	Incubate for 10–20 minutes* at room temperature (15–30°C) in the dark .
10.	Add 100 µl stop solution (STOP) into each well and mix well.
11.	Determine absorption immediately with an ELISA reader at 450 nm against 620 nm (or 690 nm) as a reference. If no reference wavelength is available, read only at 450 nm. If the extinction of the highest standard exceeds the range of the photometer, absorption must be measured immediately at 405 nm against 620 nm as a reference.

* The intensity of the colour change is temperature sensitive. We recommend observing the colour change and stopping the reaction upon good differentiation.

** We recommend shaking the strips at 550 rpm with an orbit of 2 mm.

8. RESULTS

For result evaluation, please use a four parametric logit-log model based on the standard curve of the respective kit lot and the calibrator value (CAL). All essential information on the standard curve is provided on the QC data sheet of the respective product lot.

The calibration curve can be expressed either by the concentration of each standard with its corresponding optical density or by the four parameters A,B,C and D. In both cases the optical density of the calibrator (CAL) is essential. Depending on your evaluation software program, either the one or the other kind of data described above should be entered.

Caution: Please make sure that all parameters and values are transferred accurately into your software as minor deviations can cause severe errors during evaluation.

The plausibility of the duplicate values should be examined before the automatic evaluation of the results. If this option is not available with the programme used, the duplicate values should be evaluated manually.

Serum and plasma

The obtained results have to be multiplied by the **dilution factor of 100 or 500** to get the actual concentrations.

If samples were diluted **1:4 000** or **1:8 000**, the obtained results have to be multiplied by the **dilution factor of 4 000 or 8 000** respectively.

Dried blood spots

The obtained results have to be multiplied by the **dilution factor of 60** to get the actual concentrations.

Urine

The obtained results have to be multiplied by the **dilution factor of 5** to get the actual concentrations.

Stool

The obtained results have to be multiplied by the **dilution factor of 50** to get the actual concentrations.

In case **another dilution factor** has been used, multiply the obtained result by the dilution factor used.

9. LIMITATIONS

Samples with concentrations above the measurement range can be further diluted and re-assayed. Please consider this higher dilution when calculating the results.

Samples with concentrations lower than the measurement range cannot be clearly quantified. The upper limit of the measurement range can be calculated as:

highest concentration of the standard curve × sample dilution factor to be used

The lower limit of the measurement range can be calculated as:

Analytical sensitivity × sample dilution factor to be used

Analytical sensitivity see chapter "Performance Characteristics".

10. QUALITY CONTROL

Immundiagnostik AG recommends the use of external controls for internal quality control, if possible.

Control samples should be analysed with each run. Results, generated from the analysis of control samples, should be evaluated for acceptability using appropriate statistical methods. The results for the samples may not be valid if within the same assay one or more values of the quality control sample are outside the acceptable limits.

Reference range

We recommend each laboratory to establish its own reference range.

11. PERFORMANCE CHARACTERISTICS

Analytical Sensitivity

The zero-standard was measured 18 times. The detection limit was set as $B_0 + 2 SD$ and estimated to be 0,921 ng/ml.

Precision and reproducibility

Intra-Assay (n = 20)

The reproducibility of two results in one measurement series was evaluated. Two samples were analysed 20 times by one person using the CRP ELISA.

Sample	CRP [ng/ml]	CV [%]
1	23.3	6
2	99.4	5.5

Inter-Assay (n = 15)

The reproducibility of two results at different days was evaluated. Two samples were analysed at different days over a period of three months by three different persons using the CRP ELISA.

Sample	CRP [ng/ml]	CV [%]
1	22.1	11.6
2	90.4	13.8

12. PRECAUTIONS

- All reagents in the kit package are for research use only.
- Human materials used in kit components were tested and found to be negative for HIV, Hepatitis B and Hepatitis C. However, for safety reasons, all kit components should be treated as potentially infectious.
- Kit reagents contain sodium azide or ProClin as bactericides. Sodium azide and ProClin are toxic. Substrates for the enzymatic color reactions are toxic and carcinogenic. Avoid contact with skin or mucous membranes.
- The stop solution consists of diluted sulphuric acid, a strong acid. Although diluted, it still must be handled with care. It can cause burns and should be handled with gloves, eye protection, and appropriate protective clothing. Any

spill should be wiped up immediately with copious quantities of water. Do not breath vapour and avoid inhalation.

13. TECHNICAL HINTS

- Do not interchange different lot numbers of any kit component within the same assay. Furthermore we recommend not assembling wells of different microtiter plates for analysis, even if they are of the same batch.
- Control samples should be analysed with each run.
- Reagents should not be used beyond the expiration date stated on kit label.
- Substrate solution should remain colourless until use.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- Avoid foaming when mixing reagents.
- Do not mix plugs and caps from different reagents.
- The assay should always be performed according to the enclosed manual.

14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE












- The guidelines for laboratories should be followed.
- Incubation time, incubation temperature and pipetting volumes of the components are defined by the producer. Any variation of the test procedure, which is not coordinated with the producer, may influence the results of the test. Immundiagnostik AG can therefore not be held responsible for any damage resulting from incorrect use.
- Warranty claims and complaints regarding deficiencies must be logged within 14 days after receipt of the product. The product should be send to Immundiagnostik AG along with a written complaint.

15. REFERENCES

1. Doumatey, Ayo P, Jie Zhou, Adebowale Adeyemo, and Charles Rotimi. 2014. "High Sensitivity C-Reactive Protein (Hs-CRP) Remains Highly Stable in Long-Term Archived Human Serum." *Clinical Biochemistry* **47** (4-5) (March): 315–8. doi:10.1016/j.clinbiochem.2013.12.014.

2. Koenig, Wolfgang, Hannelore Löwel, Jens Baumert, and Christa Meisinger. 2004. "C-Reactive Protein Modulates Risk Prediction Based on the Framingham Score: Implications for Future Risk Assessment: Results from a Large Cohort Study in Southern Germany." *Circulation* **109** (11) (March 23): 1349–53. doi:10.1161/01.CIR.0000120707.98922.E3.
3. Pearson, T. a. 2003. "American Heart Association Guide for Improving Cardiovascular Health at the Community Level: A Statement for Public Health Practitioners, Healthcare Providers, and Health Policy Makers From the American Heart Association Expert Panel on Population and Pre." *Circulation* **107** (4) (February 4): 645–651. doi:10.1161/01.CIR.0000054482.38437.13.
4. Ridker, P M, C H Hennekens, J E Buring, and N Rifai. 2000. "C-Reactive Protein and Other Markers of Inflammation in the Prediction of Cardiovascular Disease in Women." *The New England Journal of Medicine* **342** (12) (March 23): 836–43. doi:10.1056/NEJM200003233421202.
5. Salzer, Jonatan, Göran Hallmans, Maria Nyström, Hans Stenlund, Göran Wadell, and Peter Sundström. 2013. "Vitamin A and Systemic Inflammation as Protective Factors in Multiple Sclerosis." *Multiple Sclerosis (Houndmills, Basingstoke, England)* **19** (8) (July 18): 1046–51. doi:10.1177/1352458512472752.

Used symbols:

	Temperature limitation		Catalogue Number
	For research use only		To be used with
	Manufacturer		Contains sufficient for <cn> tests
	Lot number		Use by
	Attention		Consult instructions for use
	Consult specification data sheet		