

PermaFluor™ Aqueous Mounting Medium

Please note this data sheet has been changed effective December 19, 2011

INTENDED USE

For In Vitro Diagnostic Use

<u>AVAILABILITY:</u>	<u>Catalog #</u>	<u>Volume</u>
	TA-006-FM	6 ml
	TA-030-FM	30ml

<u>SPECIFICITY:</u>	N/A
<u>ENZYME:</u>	N/A
<u>CHROMOGEN/SUBSTRATE:</u>	N/A

DESCRIPTION

Immobilizes coverglass over section, enhances FITC intensity and reduces fading to a significant degree.

WARNINGS & PRECAUTIONS

Reagent should be at room temperature before opening vial for use. **DO NOT FREEZE.**
NO MSDS IS REQUIRED FOR THIS PRODUCT. The only chemical of a toxic nature in this product is sodium azide, which is present at a concentration below 0.1%.

STORAGE & SHELF LIFE

Store at 2-8°C.

MICROBIOLOGICAL STATE

Product(s) not sterile.

MATERIALS REQUIRED BUT NOT PROVIDED

N/A

SPECIMEN & REAGENT PREPARATION

Refer to Procedure.

PROCEDURE**FOR USE IN FLUORESCENCE MICROSCOPY**


After completion of immunostaining, remove slide from water, apply sufficient drops of PermaFluor Aqueous Mounting Medium upon section and mount coverglass. Coverslip should be immobilized within 30 minutes at room temperature. To avoid air bubbles within mounting medium, DO NOT SHAKE VIAL. If air bubbles are evident during coverslipping, carefully trim the vial tip with a sharp blade to create a larger opening and improve media flow. These directions are recommendations; however, they may be altered according to procedures established in-house.

REFERENCES

N/A

TROUBLESHOOTING

Lab Vision Corporation
46360 Fremont Blvd.
Fremont, CA 94538-6406, USA
US Toll Free: 1 (800) 522-7270
Phone: +1 (269) 544-5600
Fax: +1 (269) 372-2674
www.thermoscientific.com/labvision



Thermo Fisher Scientific
Anatomical Pathology
Tudor Road, Manor Park
Runcorn, Cheshire WA7 1TA, UK
Tel: +44 (0) 1928 534 050
Fax: +44 (0) 1928 534 049
sales.ap.uk@thermofisher.com

Please contact Thermo Fisher Scientific Technical Support by phone (1-269-544-5600 or 1-800-522-7270) or by email (lab.reagents@thermofisher.com).



Lab Vision Corporation
46360 Fremont Blvd.
Fremont, CA 94538-6406, USA
US Toll Free: 1 (800) 522-7270
Phone: +1 (269) 544-5600
Fax: +1 (269) 372-2674
www.thermoscientific.com/labvision



IVD

EC REP

Thermo Fisher Scientific
Anatomical Pathology
Tudor Road, Manor Park
Runcorn, Cheshire WA7 1TA, UK
Tel: +44 (0) 1928 534 050
Fax: +44 (0) 1928 534 049
sales.ap.uk@thermofisher.com