



## Colorimetric Microplate Assay for Total Glutathione (tGSH)

Product Number: GT 20

Store Components per Label.

FOR RESEARCH USE ONLY

Document # 070420

Page 1 of 5

# Colorimetric Microplate Assay for Total Glutathione (tGSH)


For Research Use Only

## INTRODUCTION

Glutathione (γ-glutamylcysteinylglycine or GSH) is a naturally occurring tripeptide whose nucleophilic and reducing properties play a central role in metabolic pathways, as well as in the antioxidant system of most aerobic cells. GSH is required as a coenzyme by a variety of enzymes including glutathione peroxidase, glutathione S-transferase and thioltransferase. GSH also plays a major role in drug metabolism, calcium metabolism, the γ-glutamyl cycle, cell membrane and blood platelet functions. GSH is crucial to a variety of life processes, including the detoxification of xenobiotics, maintenance of the -SH level of proteins, thiol-disulfide exchange, removal of hydroperoxides and free radicals, and amino acid transport across membranes. Physiological values for the concentration of intracellular GSH generally range from 1 to 10 mM. Although many methods have been described for the assay of GSH, most of the reliable ones have been labor intensive.

## PRINCIPLES OF PROCEDURE

This kit employs a kinetic enzymatic recycling assay based on the oxidation of GSH by 5,5'-dithiobis(2-nitrobenzoic acid) [DTNB] to measure the total glutathione (tGSH) content of biological samples. The included glutathione standards or treated samples are added to the microtiter plate wells, followed by DTNB and glutathione reductase. Addition of β-NADPH<sub>2</sub> to the wells initiates the progressive reduction of DTNB by GSH, causing a color increase that is monitored at 405 nm. The rate of color change, followed over a 5 minute time period, is proportional to the tGSH concentration. Consequently, the concentration of tGSH in unknown samples may be determined by reference to the standard curve. GSH reacts with DTNB to produce a colored ion, which absorbs light at 405 nm, and a mixed disulphide. This disulphide reacts with further quantities of GSH present to liberate another ion and GSSG. GSSG is reduced enzymatically to GSH which then re-enters the cycle (see Figure 1). Since GSSG represents only a small percentage of total acid-solution free glutathione, the resulting values for tGSH (which encompasses both GSH and GSSG) are expressed in units of GSH equivalents.



---

## MATERIALS PROVIDED

---

1. Glutathione Standard (reduced form GSH); 2 mM; 100  $\mu$ L; Store at -20°C
2. 5,5'-dithiobis(2-nitrobenzoic acid) (DTNB); 3 mg; Store at room temperature
3.  $\beta$ -Nicotinamide Adenine Dinucleotide Phosphate, reduced form ( $\beta$ -NADPH<sub>2</sub>); 3.6 mg; Store at -20°C
4. Glutathione Oxidoreductase; 25  $\mu$ L; Store at 4°C
5. Assay Buffer; 100 mL; pH 7.5
6. Metaphosphoric Acid (MPA); 60 mL; 5% (50 g/L)
7. 96 Well Microplate
8. Kit Instruction

---

## MATERIALS NEEDED BUT NOT PROVIDED

---

1. Plate reader with kinetic reading capability and 405 nm or 412 nm filters
2. Adjustable multi-channel and single-channel pipettors with disposable tips
3. Disposable glass test tubes
4. Vortex
5. Microcentrifuge tubes and 15 mL polypropylene centrifuge tubes

---

## STORAGE

---

Store kit and all of its components according to the label specifications at all times.

---

## WARNINGS AND PRECAUTIONS

---

1. Use aseptic technique when opening and dispensing reagents.
2. This kit is designed to work properly as provided and instructed. Additions, deletions or substitutions to the procedure or reagents are not recommended, as it may be detrimental to the assay.
3. Exercise universal precautions during the performance or handling of this kit or any component contained therein.

---

## PROCEDURAL NOTES

---

1. Reagents can be used immediately upon removal from refrigeration.
2. Use fresh pipette tips when transferring or pipetting reagents from stock reagents.
3. Pipette tips should be pre-wetted or rinsed prior to dispensing reagent. To do this, adjust pipette to the volume desired and draw up reagent of choice into the tip 2-3 times before uptake for dispensing.
4. To minimize error due to handling, wipe the exterior bottom of the microplate wells with a lint-free paper towel.

---

## SAMPLE COLLECTION, PREPARATION, AND STORAGE

---

### Whole Blood

1. Immediately after collection; the blood samples ( $\geq$  0.25 mL) are deproteinized by the addition of four volumes of ice-cold 5% metaphosphoric acid (MPA).
2. Thoroughly mix and centrifuge at 13,000 x g at 4°C for 2 minutes. If the samples are not going to be processed further at this time, immediately store at -70°C. Process within 3 weeks.
3. Collect the upper aqueous layer and keep at 0-4°C for assay within 24 hours. If the experiment will not be performed within 24 hours, the MPA extract should be aliquoted into vials and stored at -20°C.
4. Just before analysis, the MPA extracts should be diluted 1:20 with assay buffer.

**Erythrocyte Lysate**

1. Immediately after collection, centrifuge at least 0.5 mL of whole blood at 3,000 x g at 4°C for 5 minutes.
2. Discard the plasma supernatant. If the sample will not be assayed immediately, store the erythrocyte pellet at -70°C and finish the processing within 3 weeks.
3. Resuspend the erythrocyte pellet in four volumes of ice-cold 5% MPA. This should be done on ice.
4. Thoroughly mix and leave on ice for 15 minutes. Centrifuge at 3,000 x g at 4°C for 10 minutes.
5. Collect the upper clear aqueous layer and keep at 4°C. Assay within 4 hours.

**Hepatocyte Lysate**

1. Resuspend the hepatocyte pellet (~10<sup>6</sup>) in 0.5 mL of ice-cold 5% MPA solution.
2. Homogenize or sonicate the cell suspension and leave on ice for 5 minutes.
3. Centrifuge the homogenate at 3,000 x g at 4°C for 10 minutes.
4. Collect the upper clear aqueous layer and keep at 0-4°C. Test within 4 hours.
5. Just before analysis, the supernatant should be diluted 1:10 with assay buffer.

**Tissue Extract**

1. Wash tissue in 1x phosphate buffered saline (PBS), pH 7.2
2. Blot the tissue on filter paper.
3. Place the tissue (20 mg to 1 g) in 10 mL of ice-cold 5% MPA solution.
4. Homogenize thoroughly.
5. Centrifuge homogenate at 3,000 x g at 4°C for 10 minutes.
6. Collect the upper clear aqueous layer and keep at 0-4°C. Assay within 4 hours.

**Human Urine and Saliva**

1. Immediately after collecting the urine or saliva, add 4 volumes of ice-cold 5% MPA to the sample.
2. Thoroughly mix and centrifuge at 13,000 x g at 4°C for 2 minutes.
3. Collect the upper clear aqueous layer and keep at 0-4°C. Assay within 4 hours.

**Cell Cultures in 96 well plates**

1. After incubation, the cells (x10<sup>4</sup>) are washed gently with phosphate buffered saline (PBS), pH 7.2.
2. Aspirate the washing solution and add 55-100 µL of 5% MPA to each well.
3. The culture plates are then frozen at -80°C and subsequently thawed at 37°C. This freeze/thaw cycle should be done twice and the plate left on ice for 15 minutes.
4. Remove 50 µL of the MPA extract from each well and transfer it to the assay microtiter plate using a multiple-channel pipettor.
5. Transfer all the wells such that the plate configuration of the treated wells will be maintained in the assay microtiter plate. Leave space to run the GSH standard curve.

---

**REAGENT PREPARATION**

---

**NOTE:** The stability of reagents, once reconstituted, is greatly diminished. The kit should be used in its entirety within 2 hours of reagent reconstitution. During this time reagents should be stored as recommended in the following reconstitution instruction.

1. Add 6 mL of Assay Buffer to the 3.0 mg DTNB and vortex. Leave at room temperature.
2. Add 6 mL of Assay Buffer to the 3.6 mg β-NADPH<sub>2</sub> and vortex. Leave on ice.
3. Transfer 22 µL of the Glutathione Oxidoreductase to 6 mL of Assay Buffer and vortex. Leave on ice.

---

## STANDARD CURVE PREPARATION

---

Prepare the GSH standards within a concentration range from 0.5 to 20  $\mu\text{M}$  in 5% MPA or anywhere between 0-100  $\mu\text{M}$  as needed. Prepare and label standards as follows in Table 1.

**Table 1.**

Standard	GSH Concentration $\mu\text{M}$ ( $\mu\text{g/mL}$ )	5% MPA (mL)	Transfer Volume (mL)	Transfer Source	Final Volume (mL)
S <sub>7</sub>	20 (6.15 $\mu\text{g/mL}$ )	3.96	0.04	Stock	1
S <sub>6</sub>	15 (4.61 $\mu\text{g/mL}$ )	1	3	S <sub>7</sub>	1.33
S <sub>5</sub>	10 (3.07 $\mu\text{g/mL}$ )	1.33	2.67	S <sub>6</sub>	2
S <sub>4</sub>	5.0 (1.54 $\mu\text{g/mL}$ )	2	2	S <sub>5</sub>	2
S <sub>3</sub>	2.5 (0.77 $\mu\text{g/mL}$ )	2	2	S <sub>4</sub>	2.4
S <sub>2</sub>	1.0 (0.31 $\mu\text{g/mL}$ )	2.4	1.6	S <sub>3</sub>	2
S <sub>1</sub>	0.5 (0.15 $\mu\text{g/mL}$ )	2	2	S <sub>2</sub>	4
B <sub>0</sub>	0 (0 $\mu\text{g/mL}$ )	2	---	---	2

---

## ASSAY PROCEDURE

---

1. Add 50  $\mu\text{L}$  of each GSH standard or diluted or undiluted sample extract to each well of the microtiter plate. Both standards and samples are analyzed in duplicate or triplicate as needed. **NOTE:** For increased accuracy, the standards can be transferred into (mini) cluster tubes. Use an eight-channel pipettor to transfer the standards into the microtiter plate.
2. Add 50  $\mu\text{L}$  each of DTNB and Glutathione Oxidoreductase solution to each well. The use of an eight-channel pipettor is recommended.
3. Incubate the plate for 10 minutes at room temperature.
4. Add 50  $\mu\text{L}$  of  $\beta\text{-NADPH}_2$  solution to each well to start the reaction.
5. Within 1 minute after the addition of  $\beta\text{-NADPH}_2$ , read the plate using a kinetic program that can monitor the reaction at 1 minute intervals for 10 minutes. The rate of color change is monitored in the 405-412 nm range (depending on the filter wavelengths available on your plate reader). **Note:** If your plate reader does not have a kinetic reading function, then read the plate and record the OD value at 0 minutes and again at 10 minutes.

---

## CALCULATIONS


---

Plot the standard curve with Mean V versus GSH concentrations. The total GSH concentration of the samples can be calculated from the standard curve.

If a kinetic microplate reader is not available, the Mean V (the rate of color change) can be estimated by subtracting the OD value at 0 minutes from the OD value at 10 minutes and dividing the result by 10. That is:

$$\text{Mean V} = (\text{OD at 10 minutes} - \text{OD at 0 minutes}) / 10$$

The rate of color change represented as Mean V is proportional to the GSH standard concentrations.

**Sample Curves:****DISCLAIMER**

This information is believed to be correct but does not purport to be all-inclusive and shall be used only as a guide. Oxford Biomedical Research, Inc. shall not be held liable for any damage resulting from handling or from contact with the above product. See catalog for additional terms and conditions of sale.

**ORDERING INFORMATION**

For additional kits or a complete catalog please call 1-800-692-4633.

**TECHNICAL SUPPORT**

If you need technical information or assistance with assay procedures, call our Technical Support Department at 1-800-692-4633 or 1-248-852-8815. Our staff will be happy to answer your questions about this or any other product in the Oxford Biomedical line.

**GUARANTEE AND LIMITATION OF REMEDY**

Oxford Biomedical Research, Inc. makes no guarantee of any kind, expressed or implied, which extends beyond the description of the material in this kit, except that these materials and this kit will meet our specifications at the time of delivery. Buyer's remedy and Oxford Biomedical Research, Inc.'s sole liability hereunder is limited to, at Oxford Biomedical Research, Inc.'s option, refund of the purchase price of, or the replacement of, material that does not meet our specification. By acceptance of our products, Buyer indemnifies and holds Oxford Biomedical Research, Inc. harmless against, assumes all liability for the consequence of its use or misuse by the Buyer, its employees, or others. Said refund or replacement is conditioned of Buyer notifying Oxford Biomedical Research, Inc. within (30) days of the receipt of product. Failure of Buyer to give said notice within thirty (30) days shall constitute a waiver by the Buyer of all claims hereunder with respect to said material(s).

Oxford Biomedical Research, Inc.  
P.O. Box 522  
Oxford, MI 48371 U.S.A.

Orders: 800-692-4633  
Technical Service: 248-852-8815  
Fax: 248-852-4466  
E-mail: [info@oxfordbiomed.com](mailto:info@oxfordbiomed.com)

Made in the U.S.A.