

# Quantikine<sup>®</sup> ELISA

## Human Myeloperoxidase Immunoassay

Catalog Number DMYE00B

For the quantitative determination of human Myeloperoxidase (MPO) concentrations in cell culture supernates, cell lysates, serum, platelet-poor plasma, saliva, and urine.

This package insert must be read in its entirety before using this product.  
For research use only. Not for use in diagnostic procedures.

# TABLE OF CONTENTS

SECTION	PAGE
INTRODUCTION .....	1
PRINCIPLE OF THE ASSAY.....	2
LIMITATIONS OF THE PROCEDURE .....	2
TECHNICAL HINTS.....	2
MATERIALS PROVIDED & STORAGE CONDITIONS .....	3
OTHER SUPPLIES REQUIRED .....	3
PRECAUTIONS.....	4
SAMPLE COLLECTION & STORAGE.....	4
CELL LYSIS PROCEDURE.....	5
SAMPLE PREPARATION.....	5
REAGENT PREPARATION .....	6
ASSAY PROCEDURE .....	7
CALCULATION OF RESULTS.....	8
TYPICAL DATA.....	8
PRECISION .....	9
RECOVERY.....	9
LINEARITY.....	10
SENSITIVITY .....	10
CALIBRATION .....	10
SAMPLE VALUES.....	11
SPECIFICITY.....	11
REFERENCES .....	12
PLATE LAYOUT .....	13

## MANUFACTURED AND DISTRIBUTED BY:

### USA & Canada | R&D Systems, Inc.

614 McKinley Place NE, Minneapolis, MN 55413, USA  
TEL: (800) 343-7475 (612) 379-2956 FAX: (612) 656-4400  
E-MAIL: info@RnDSystems.com

## DISTRIBUTED BY:

### UK & Europe | R&D Systems Europe, Ltd.

19 Barton Lane, Abingdon Science Park, Abingdon OX14 3NB, UK  
TEL: +44 (0)1235 529449 FAX: +44 (0)1235 533420  
E-MAIL: info@RnDSystems.co.uk

### China | R&D Systems China Co., Ltd.

24A1 Hua Min Empire Plaza, 726 West Yan An Road, Shanghai PRC 200050  
TEL: +86 (21) 52380373 FAX: +86 (21) 52371001  
E-MAIL: info@RnDSystemsChina.com.cn

## INTRODUCTION

Myeloperoxidase (MPO) is a heme-containing enzyme belonging to the XPO subfamily of peroxidases. It is an abundant neutrophil and monocyte glycoprotein that catalyzes the hydrogen peroxide dependent formation of hypochlorous acid (HOCl) and other reactive species (1, 2). Reaction of these compounds with macromolecules results in the nitrosylation, chlorination, and oxidation of tyrosine residues, lipids, and cholesterol, and the intermolecular crosslinking of proteins and DNA (1, 3-8). MPO is synthesized as a preproprotein that is proteolytically processed to remove a 48 amino acid (aa) signal peptide, a 116 aa propeptide, the C-terminal serine, and a 6 aa internal peptide which generates separate 60 kDa heavy and 12 kDa light chains (9). Other post-translational modifications of MPO include insertion of a heme moiety, glycosylation, and phosphorylation of mannose residues (10, 11). Enzymatically active MPO is a disulfide-linked tetramer that contains two heme groups and two copies each of the heavy and light chains (10, 11). Alternate splicing results in two additional isoforms of MPO, one with a 32 aa insertion in the light chain, and another with a deletion of the signal sequence and 47 aa of the propeptide (9). Mature human MPO shares 87-88% aa sequence identity with canine, mouse, and rat MPO. It shares 71%, 56%, and 47% aa sequence identity with comparable regions of human eosinophil peroxidase, lactoperoxidase, and thyroid peroxidase, respectively.

MPO binds albumin, the macrophage mannose receptor, cytokeratin 1 on vascular endothelial cells, high molecular weight kininogen, and the integrin CD11b/CD18 on neutrophils (12-15). These interactions promote MPO clearance, a reduction of nitric oxide and bradykinin levels, reduced vasodilation, and continued neutrophil activation (12-16). MPO gene expression is under the control of thyroid hormone, retinoic acid, estrogen, and PPAR $\gamma$  receptors (17, 18). A polymorphism within the MPO gene promoter confers increased steroid responsiveness and is associated with the development or severity of a variety of diseases (1, 17, 18).

Neutrophil MPO is stored in cytoplasmic azurophilic granules (19, 20). Upon cellular activation and degranulation, MPO is delivered into phagosomes where it is required for the killing of phagocytosed bacteria (10, 21). Activated neutrophils also release granule contents extracellularly. Elevated plasma MPO levels have been associated with a variety of clinical conditions including systemic inflammation, eclampsia, risk of cardiovascular events, vascular endothelial dysfunction, severity of multiple sclerosis, and prospective mortality and oxidative stress during hemodialysis (22-30). MPO levels are also elevated in arthritic synovial fluid and gingivitis crevicular fluid (31, 32). MPO is incorporated into atherosclerotic plaques and renal stones and serves as an indicator of neutrophil infiltration of inflamed or injured tissues (33-35).

The Quantikine Human MPO Immunoassay is a 4.5 hour solid-phase ELISA designed to measure human MPO in cell culture supernates, cell lysates, serum, platelet-poor plasma, saliva, and urine.

## PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for MPO has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any MPO present is bound by the immobilized antibody. After washing away any unbound substances, an enzyme-linked polyclonal antibody specific for MPO is added to the wells. Following a wash to remove any unbound antibody-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MPO bound in the initial step. The color development is stopped and the intensity of the color is measured.

## LIMITATIONS OF THE PROCEDURE

- FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.
- The kit should not be used beyond the expiration date on the kit label.
- Do not mix or substitute reagents with those from other lots or sources.
- If samples generate values higher than the highest standard, further dilute the samples with the appropriate Calibrator Diluent and repeat the assay.
- Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.
- Variations in sample collection, processing, and storage may cause sample value differences.
- This assay is designed to eliminate interference by enzymes, proteins, and other factors present in biological samples. Until all factors have been tested in the Quantikine Immunoassay, the possibility of interference cannot be excluded.

## TECHNICAL HINTS

- When mixing or reconstituting protein solutions, always avoid foaming.
- To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- When using an automated plate washer, adding a 30 second soak period following the addition of Wash Buffer, and/or rotating the plate 180 degrees between wash steps may improve assay precision.
- Substrate Solution should remain colorless until added to the plate. Keep Substrate Solution protected from light. Substrate Solution should change from colorless to gradations of blue.
- Stop Solution should be added to the plate in the same order as the Substrate Solution. The color developed in the wells will turn from blue to yellow upon addition of the Stop Solution. Wells that are green in color indicate that the Stop Solution has not mixed thoroughly with the Substrate Solution.

## MATERIALS PROVIDED & STORAGE CONDITIONS

Store the unopened kit at 2-8 °C. Do not use past kit expiration date.

PART	PART #	DESCRIPTION	STORAGE OF OPENED/ RECONSTITUTED MATERIAL
MPO Microplate	893941	96 well polystyrene microplate (12 strips of 8 wells) coated with a rat monoclonal antibody against MPO.	Return unused wells to the foil pouch containing the desiccant pack. Reseal along entire edge of zip-seal. May be stored for up to 1 month at 2-8 °C.*
MPO Conjugate	893942	21 mL of polyclonal antibody against MPO conjugated to horseradish peroxidase with preservatives.	May be stored for up to 1 month at 2-8 °C.*
MPO Standard	893943	100 ng of natural human MPO in a buffered protein base with preservatives; lyophilized.	
Assay Diluent RD1-27	895245	11 mL of a buffered protein base with preservatives.	
Calibrator Diluent RD5K	895119	21 mL of a buffered protein base with preservatives. <i>For cell culture supernate/cell lysate/saliva samples.</i>	
Calibrator Diluent RD6-58	895951	21 mL of a buffered protein base with preservatives. <i>For serum/plasma/urine samples.</i>	
Wash Buffer Concentrate	895003	21 mL of a 25-fold concentrated solution of buffered surfactant with preservatives.	
Color Reagent A	895000	12 mL of stabilized hydrogen peroxide.	
Color Reagent B	895001	12 mL of stabilized chromogen (tetramethylbenzidine).	
Stop Solution	895032	6 mL of 2 N sulfuric acid.	
Plate Sealers	N/A	4 adhesive strips.	

\* Provided this is within the expiration date of the kit.

## OTHER SUPPLIES REQUIRED

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 500 mL graduated cylinder.
- Horizontal orbital microplate shaker (0.12" orbit) capable of maintaining a speed of 500 ± 50 rpm.
- Test tubes for dilution of standards and samples.

### If using cell lysate samples, the following is also required:

- Cell Lysis Buffer 3 (R&D Systems, Catalog # 895366).
- PBS

## PRECAUTIONS

MPO is detectable in saliva. Take precautionary measures to prevent contamination of kit reagents while running this assay.

The MPO Standard provided in this kit was derived from human blood. The source material was tested at the donor level using FDA licensed methods and found to be non-reactive for anti-HIV-1/2, anti-HCV, and Hepatitis B surface antigen. As no testing can offer complete assurance of freedom from infectious agents, the Standard should be handled as if capable of transmitting infection.

The Stop Solution provided with this kit is an acid solution. Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling.

## SAMPLE COLLECTION & STORAGE

**The sample collection and storage conditions listed below are intended as general guidelines. Sample stability has not been evaluated.**

**Cell Culture Supernates** - Remove particulates by centrifugation and assay immediately or aliquot and store samples at  $\leq -20$  °C. Avoid repeated freeze-thaw cycles.

**Cell Lysates** - Cells must be lysed before assaying. Refer to the Cell Lysis Procedure.

**Serum** - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at  $\leq -20$  °C. Avoid repeated freeze-thaw cycles.

**Platelet-poor Plasma** - Collect plasma on ice using EDTA or heparin as an anticoagulant. Centrifuge at 2-8 °C for 15 minutes at 1000 x g within 30 minutes of collection. An additional centrifugation step of the separated plasma at 10,000 x g for 10 minutes at 2-8 °C is recommended for complete platelet removal. Assay immediately or aliquot and store samples at  $\leq -20$  °C. Avoid repeated freeze-thaw cycles.

**Note:** *Citrate plasma has not been validated for use in this assay.*

**MPO is present in neutrophil granules and is released upon neutrophil exposure to activated platelets. Therefore, to measure circulating levels of MPO, platelet-free plasma should be collected for measurement. It should be noted that many protocols for plasma preparation, including procedures recommended by the Clinical Laboratory and Standards Institute (CLSI), result in incomplete removal of platelets from blood. This will cause variable and irreproducible results for assays of factors released by platelet activation.**

**Saliva** - Collect saliva in a tube and centrifuge for 5 minutes at 10,000 x g. Collect the aqueous layer, and assay immediately or aliquot and store samples at  $\leq -20$  °C. Avoid repeated freeze-thaw cycles.

**Urine** - Aseptically collect the first urine of the day (mid-stream), voided directly into a sterile container. Centrifuge to remove particulate matter, and assay immediately or aliquot and store at  $\leq -20$  °C. Avoid repeated freeze-thaw cycles.

## CELL LYSIS PROCEDURE

Use the following procedure for the preparation of cell lysate samples.

1. Perform a 5-fold dilution of Cell Lysis Buffer 3 (available from R&D Systems) with deionized or distilled water.
2. Wash cells one time in cold PBS.
3. Resuspend cells at  $5 \times 10^6$  cells/mL in diluted Cell Lysis Buffer 3.
4. Incubate with gentle agitation for 30 minutes at room temperature and freeze/thaw cells once at  $\leq -20$  °C.
5. Centrifuge to remove cell debris.
6. Assay immediately or aliquot the lysis supernates and store at  $\leq -20$  °C until ready for use.

## SAMPLE PREPARATION

Serum samples require a 50-fold dilution. A suggested 50-fold dilution is 10  $\mu$ L of sample + 90  $\mu$ L of Calibrator Diluent RD6-58 followed by 50  $\mu$ L of diluted sample + 200  $\mu$ L of Calibrator Diluent RD6-58.

Platelet-poor plasma samples require a 10-fold dilution. A suggested 10-fold dilution is 20  $\mu$ L of sample + 180  $\mu$ L of Calibrator Diluent RD6-58.

Saliva samples require a 100-fold dilution. A suggested 100-fold dilution is 20  $\mu$ L of sample + 180  $\mu$ L of Calibrator Diluent RD5K followed by 20  $\mu$ L of diluted sample + 180  $\mu$ L of Calibrator Diluent RD5K.

## REAGENT PREPARATION

**Bring all reagents to room temperature before use.**


**Note:** MPO is detectable in saliva. It is recommended that a face mask and gloves be used to protect kit reagents from contamination.

**Wash Buffer** - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 mL of Wash Buffer Concentrate into deionized or distilled water to prepare 500 mL of Wash Buffer.

**Substrate Solution** - Color Reagents A and B should be mixed together in equal volumes within 15 minutes of use. Protect from light. 200  $\mu$ L of the resultant mixture is required per well.

**MPO Standard** - Reconstitute the MPO Standard with 1.0 mL of deionized or distilled water. This reconstitution produces a stock solution of 100 ng/mL. Mix the standard to ensure complete reconstitution and allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions.

Pipette 900  $\mu$ L of Calibrator Diluent RD5K (for cell culture supernate/cell lysate/saliva samples) or Calibrator Diluent RD6-58 (for serum/plasma/urine samples) into the 10 ng/mL tube. Pipette 500  $\mu$ L of the appropriate Calibrator Diluent into the remaining tubes. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 10 ng/mL standard serves as the high standard. The appropriate Calibrator Diluent serves as the zero standard (0 ng/mL).





## ASSAY PROCEDURE

**Bring all reagents and samples to room temperature before use. It is recommended that all standards, samples, and controls be assayed in duplicate.**

**Note:** *MPO is detectable in saliva. It is recommended that a face mask and gloves be used to protect kit reagents from contamination.*

1. Prepare all reagents, working standards, and samples as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the foil pouch containing the desiccant pack, and reseal.
3. Add 100  $\mu$ L of Assay Diluent RD1-27 to each well.
4. Add 50  $\mu$ L of Standard, control, or sample\* per well. Cover with the adhesive strip provided. Incubate for 2 hours at room temperature on a horizontal orbital microplate shaker (0.12" orbit) set at  $500 \pm 50$  rpm. A plate layout is provided to record standards and samples assayed.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with Wash Buffer (400  $\mu$ L) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 200  $\mu$ L of MPO Conjugate to each well. Cover with a new adhesive strip. Incubate for 2 hours at room temperature on the shaker.
7. Repeat the aspiration/wash as in step 5.
8. Add 200  $\mu$ L of Substrate Solution to each well. Incubate for 30 minutes at room temperature **on the benchtop. Protect from light.**
9. Add 50  $\mu$ L of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green or the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
10. Determine the optical density of each well within 30 minutes, using a microplate reader set to 450 nm. If wavelength correction is available, set to 540 nm or 570 nm. If wavelength correction is not available, subtract readings at 540 nm or 570 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Readings made directly at 450 nm without correction may be higher and less accurate.

\*Samples may require dilution. See Sample Preparation section.

## CALCULATION OF RESULTS

Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density.


Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the MPO concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

Since samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

## TYPICAL DATA


These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.

### CALIBRATOR DILUENT RD5K



(ng/mL)	O.D.	Average	Corrected
0	0.008 0.009	0.009	—
0.156	0.049 0.049	0.049	0.040
0.313	0.094 0.094	0.094	0.085
0.625	0.186 0.187	0.187	0.178
1.25	0.371 0.373	0.372	0.363
2.5	0.719 0.734	0.727	0.718
5	1.344 1.369	1.357	1.348
10	2.304 2.401	2.353	2.344

### CALIBRATOR DILUENT RD6-58



(ng/mL)	O.D.	Average	Corrected
0	0.013 0.013	0.013	—
0.156	0.049 0.051	0.050	0.037
0.313	0.091 0.096	0.094	0.081
0.625	0.172 0.173	0.173	0.160
1.25	0.336 0.339	0.338	0.325
2.5	0.617 0.624	0.621	0.608
5	1.171 1.175	1.173	1.160
10	2.134 2.202	2.168	2.155

## PRECISION

### Intra-assay Precision (Precision within an assay)

Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.

### Inter-assay Precision (Precision between assays)

Three samples of known concentration were tested in twenty separate assays to assess inter-assay precision.

## CELL CULTURE SUPERNATE/CELL LYSATE/SALIVA ASSAY

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean (ng/mL)	0.994	3.36	6.68	0.974	2.99	6.23
Standard deviation	0.034	0.085	0.123	0.078	0.18	0.38
CV (%)	3.4	2.5	1.8	8.0	6.0	6.0

## SERUM/PLASMA/URINE ASSAY

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean (ng/mL)	1.03	3.13	6.63	1.07	3.29	6.71
Standard deviation	0.024	0.046	0.172	0.12	0.26	0.55
CV (%)	2.3	1.5	2.6	10.8	8.0	8.2

## RECOVERY

The recovery of MPO spiked to levels throughout the range of the assay in various matrices was evaluated.

Sample Type	Average % Recovery	Range
Cell culture media (n=4)	98	92-105%
Serum (n=4)	100	88-110%
Platelet-poor EDTA plasma (n=4)	95	90-101%
Platelet-poor heparin plasma (n=4)	91	83-100%

## LINEARITY

To assess the linearity of the assay, samples containing and/or spiked with high concentrations of MPO were serially diluted with the appropriate Calibrator Diluent to produce samples with values within the dynamic range of the assay.

		Cell culture supernates (n=4)	Serum* (n=5)	Platelet-poor		Saliva* (n=4)	Urine* (n=4)
				EDTA plasma* (n=4)	Heparin plasma* (n=4)		
1:2	Average % of Expected	98	108	106	100	102	98
	Range (%)	96-100	100-114	102-108	93-105	99-107	97-98
1:4	Average % of Expected	100	108	109	104	102	98
	Range (%)	94-103	101-117	103-115	95-111	100-105	93-102
1:8	Average % of Expected	102	110	115	106	106	99
	Range (%)	96-107	98-120	113-117	95-114	103-111	94-104
1:16	Average % of Expected	99	108	112	103	97	92
	Range (%)	95-104	100-114	101-120	90-111	97-99	85-97

\*Samples were diluted prior to assay.

## SENSITIVITY

Fifty-four assays were evaluated and the minimum detectable dose (MDD) of MPO ranged from 0.003-0.062 ng/mL. The mean MDD was 0.014 ng/mL.

The MDD was determined by adding two standard deviations to the mean optical density value of twenty zero standard replicates and calculating the corresponding concentration.

## CALIBRATION

This immunoassay is calibrated against highly purified MPO from human leukocytes.

## SAMPLE VALUES

**Serum/Plasma/Saliva/Urine** - Samples from apparently healthy volunteers were evaluated for the presence of MPO in this assay. No medical histories were available for the donors used in this study.

Sample Type	Mean (ng/mL)	Range (ng/mL)	Standard Deviation (ng/mL)
Serum (n=29)	100	21.4-229	65
Platelet-poor EDTA plasma (n=29)	7.82	2.89-27.4	4.93
Platelet-poor heparin plasma (n=29)	9.47	3.57-28.4	5.27
Saliva (n=10)	558	31.2-1428	461
Urine (n=13)	4.39	0.17-13.5	4.68

### Cell Culture Supernates -

Human monocytes were cultured in RPMI supplemented with 10% fetal bovine serum, 2 mM L-glutamine, 100 U/mL penicillin, and 100 µg/mL streptomycin sulfate. Cells were cultured unstimulated or stimulated with 1 µg/mL LPS for 1 and 6 days. Aliquots of the cell culture supernates were removed and assayed for levels of natural MPO.

Condition	Day 1 (ng/mL)	Day 6 (ng/mL)
Unstimulated	28.0	41.4
Stimulated	29.9	30.3

HL-60 human acute promyelocytic leukemia cells were cultured in RPMI supplemented with 20% fetal calf serum, 2 mM L-glutamine, 100 U/mL penicillin, and 100 µg/mL streptomycin sulfate. An aliquot of the cell culture supernate was removed, assayed for levels of natural MPO, and measured 18.2 ng/mL.

An aliquot of the same HL-60 cells was removed, and the cells were lysed at  $5 \times 10^6$  cells/mL in diluted Cell Lysis Buffer 3. The lysate was assayed for levels of natural MPO and measured 2673 ng/mL.

## SPECIFICITY

This assay recognizes natural human MPO.

The factors listed below were prepared at 100 ng/mL in Calibrator Diluent and assayed for cross-reactivity. Preparations of the following factors at 100 ng/mL in a mid-range human MPO control were assayed for interference. No significant cross-reactivity or interference was observed.

### Recombinant human:

COX1  
Glutathione Peroxidase  
HMW Kininogen  
Integrin  $\beta$ 2  
MMP-12  
Thyroid Peroxidase

### Recombinant mouse:

MPO

### Natural proteins:

bovine MPO  
human Neutrophil Elastase

Natural human Eosinophil Peroxidase cross-reacts at approximately 1.8%.

## REFERENCES

1. Klebanoff, S.J. (2005) *J. Leukoc. Biol.* **77**:598.
2. Morishita, K. *et al.* (1987) *J. Biol. Chem.* **262**:3844.
3. van Dalen, C.J. *et al.* (2000) *J. Biol. Chem.* **275**:11638.
4. Podrez, E.A. *et al.* (1999) *J. Clin. Invest.* **103**:1547.
5. Bergt, C. *et al.* (2004) *Proc. Natl. Acad. Sci.* **101**:13032.
6. Hazen, S.L. *et al.* (1996) *J. Biol. Chem.* **271**:23080.
7. Heinecke, J.W. *et al.* (1993) *J. Clin. Invest.* **91**:2866.
8. Kulcharyk, P.A. and J.W. Heinecke (2001) *Biochemistry* **40**:3648.
9. Hashinaka, K. *et al.* (1988) *Biochemistry* **27**:5906.
10. Hansson, M. *et al.* (2006) *Arch. Biochem. Biophys.* **445**:214.
11. Nauseef, W.M. *et al.* (1992) *Blood* **80**:2622.
12. Tiruppathi, C. *et al.* (2004) *Proc. Natl. Acad. Sci.* **101**:7699.
13. Shepherd, V.L. *et al.* (1990) *Am. J. Respir. Cell Mol. Biol.* **2**:335.
14. Astern, J.M. *et al.* (2007) *Am. J. Pathol.* **171**:349.
15. Lau, D. *et al.* (2005) *Proc. Natl. Acad. Sci.* **102**:431.
16. Eiserich, J.P. *et al.* (2002) *Science* **296**:2391.
17. Kumar, A.P. *et al.* (2004) *J. Biol. Chem.* **279**:8300.
18. Piedrafito, F.J. *et al.* (1996) *J. Biol. Chem.* **271**:14412.
19. Cramer, E. *et al.* (1985) *Blood* **65**:423.
20. Borregaard, N. and J.B. Cowland (1997) *Blood* **89**:3503.
21. Hirche, T.O. *et al.* (2005) *J. Immunol.* **174**:1557.
22. Sun, Y.C. *et al.* (2007) *Clin. Chim. Acta* **378**:175.
23. Andelid, K. *et al.* (2007) *Respir. Med.* **101**:888.
24. Noyan, T. *et al.* (2006) *Aust. N.Z. J. Obstet. Gynaecol.* **46**:486.
25. Meuwese, M.C. *et al.* (2007) *J. Am. Coll. Cardiol.* **50**:159.
26. Baldus, S. *et al.* (2003) *Circulation* **108**:1440.
27. Vita, J.A. *et al.* (2004) *Circulation* **110**:1134.
28. Minohara, M. *et al.* (2006) *J. Neuroimmunol.* **178**:156.
29. Kalantar-Zadeh, K. *et al.* (2006) *Am. J. Kidney Dis.* **48**:59.
30. Wu, C.C. *et al.* (2005) *Nephrol. Dial. Transplant.* **20**:1134.
31. Steinbeck, M.J. *et al.* (2007) *J. Orthop. Res.* **25**:1128.
32. Ortiz, G.C. *et al.* (1997) *Eur. J. Oral Sci.* **105**:143.
33. Daugherty, A. *et al.* (1994) *J. Clin. Invest.* **94**:437.
34. Mushtaq, S. *et al.* (2007) *Clin. Chim. Acta* **384**:41.
35. Irazuzta, J.E. *et al.* (2000) *Life Sci.* **67**:365.

**PLATE LAYOUT**

Use this plate layout to record standards and samples assayed.

12									
11									
10									
9									
8									
7									
6									
5									
4									
3									
2									
1									
	A	B	C	D	E	F	G	H	

**NOTES**