



NBT SUBSTRATE KIT FOR GLUCOSE OXIDASE

Catalog Number SK-3100

The NBT substrate kit contains all of the reagents necessary to prepare a working solution of nitroblue tetrazolium (NBT) for staining tissue sections. NBT will yield a purple/blue colored reaction product.

DISPENSING REAGENTS:

For convenience the reagents are supplied in dropper bottles. When dispensing drops, hold the bottle in an inverted vertical position and squeeze gently. The drop volumes of each reagent in this kit are not necessarily the same. Only use the drop dispensers for preparing substrate solutions. To prevent evaporation, secure the opaque caps on the bottles when they are not in use. After completion of the staining procedure, the diluted working solution should be discarded, the container rinsed with distilled water and stored together with the stock reagents in the kit box.

INSTRUCTIONS FOR USE:

Immediately before use, add two drops each of reagents 1, 2, and 3 to 5 ml of 50 mM TRIS-HCl buffer, pH 9.5, in the substrate solution mixing bottle and mix well.

Incubate tissue sections with the substrate at room temperature in the dark until suitable staining develops. Development times should be determined by the investigator but generally 10 - 30 minutes provides good staining intensity.

Wash the sections for 5 minutes in water.

Counterstain and mount, using an aqueous or non-aqueous mounting medium. (See reverse side for counterstain guidelines.)

NOTES:

If at all possible, glass-distilled water should be used in the preparation of the substrate buffer.

The reagents should be stored at 4 °C and protected from light. Reagent 2 is very light sensitive and should be kept in the closed box whenever possible.

Elevation of the incubation temperature to 37 °C will shorten the development time or may permit the use of higher dilutions of primary antibody.

In the case of non-aqueous mounting, dehydration should be carried out rapidly.

IMPORTANT: Little is known about the toxicity and carcinogenicity of the substrate kit components. Appropriate care should be exercised when using this reagent including gloves, eye protection, lab coats, and good laboratory procedures. Dispose in accordance with local regulations.